



MWH

LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

Chameleon Beverage Company, Inc.
6444 East 26th Street
Commerce, CA 90040
Attention: Hilda Menendez
Fax: 323-724-9048

Date of Issue

07/07/2011

Linda Geddes
MWH LABORATORIES



Report#: 359952

Project: ANNUAL

Group: Annual Screen

LXG: Linda Geddes

Project Manager

Laboratory certifies that the test results meet all **NELAC** requirements unless noted in the Comments section or the Case Narrative. Following the cover page are Hits Reports, Comments, QC Summary, QC Report and Regulatory Forms. This report shall not be reproduced except in full, without the written approval of the laboratory.

**MWH****LABORATORIES****STATE CERTIFICATION LIST**

State	Certification Number	State	Certification Number
Alabama	41060	Mississippi	Certified
Alaska	CA00006	Montana	Cert 0035
Arizona	AZ0455	Nevada	CA00006-2010-1
Arkansas	Certified	New Hampshire	2959-10
California – NELAP	01114CA	New Jersey	CA 008
California – ELAP	1422	New Mexico	Certified
Colorado	Certified	New York	11320
Connecticut	PH-0107	North Carolina	6701
Delaware	CA 006	North Dakota	R-009
Florida	E871024	Oregon	CA 200003-007
Georgia	947	Pennsylvania	68-565
Guam	09-006r	Rhode Island	01114CA
Hawaii	Certified	South Carolina	87016001
Idaho	Certified	South Dakota	Certified
Illinois	200033	Tennessee	TN02839
Indiana	C-CA-01	Texas	T104704230-10-1
Kansas	E-10268	Utah	Mont-1
Kentucky	90107	Vermont	VT0114
Louisiana	LA070018	Virginia	210
Maine	CA0006	Washington	C383-10a
Maryland	224	West Virginia	9943 C
Commonwealth of Northern Marianas Is.	0007;0008	Wisconsin	998316660
Massachusetts	M-CA006	Wyoming	8TMS-Q
Michigan	9906	EPA Region 5	Certified



Acknowledgement of Samples Received

Chameleon Beverage Company, Inc.

6444 East 26th Street
Commerce, CA 90040
Attn: Hilda Menendez
Phone: 323-724-8223

Customer Code: CHAMELEON
Folder #: 359952
Project: ANNUAL
Sample Group: Annual Screen
Project Manager: Linda Geddes
Phone: (626) 386-1163

The following samples were received from you on **March 29, 2011**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using MWH Laboratories.

Sample #	Sample ID	Sample Date
201103290086	Spring Water CBC 11 084 0127 EXP MAR 25 13	Mar 29, 2011 11:39
	@2378-TCDD_Dioxin @505 @515.4	
	@525 @531 @549	
	@551SODA @900 @ANIONS28	
	@ANIONS48 @COLI10 @DBP_14	
	@DBP_28 @HAA @ICP	
	@ICPMS @R226EDD @R228EDD	
	@VOA Alkalinity in CaCO3 units Anion Sum - Calculated	
	Apparent Color Asbestos by TEM - >10 microns Bicarb.Alkalinity as HCO3,calc	
	Carbon Dioxide,Free(25C)-Calc. Carbonate as CO3, Calculated Cation Sum - Calculated	
	Chloramines Chlorine Dioxide CLO41PPB	
	Cyanide Cyanide by manual distillation Endothall	
	Fluoride Free Chlorine Residual Glyphosate	
	Heterotrophic Plate Count Hydroxide as OH, Calculated Langelier Index - 25 degree	
	Mercury Odor at 60 C (TON) pH of CaCO3 saturation(25C)	
	PH, Bottled Water Phenolic Compounds-low level Specific Conductance	
	Surfactants Total Chlorine Residual Total Dissolved Solid (TDS)	
	Total Hardness as CaCO3 by ICP Turbidity @YMOLD	
	Bromate by UV/VIS	

Test Description

- @2378-TCDD_Dioxin -- 2,3,7,8-TCDD_Dioxin
- @505 -- Organochlorine Pesticides/PCBs
- @515.4 -- Chlorophenoxy Herbicides
- @525 -- Semivolatiles by GCMS
- @531 -- Aldicarbs
- @549 -- Diquat and Paraquat
- @551SODA -- EDB/DBCP/HAN by EPA 551.1
- @900 -- Gross Alpha/Beta Radiation
- @ANIONS28 -- Chloride, Sulfate by EPA 300.0
- @ANIONS48 -- Nitrate, Nitrite by EPA 300.0
- @COLI10 -- Total & Fecal Coliform, 24 Hour
- @DBP_14 -- Chlorite by 300.0



Acknowledgement of Samples Received

Chameleon Beverage Company, Inc.

6444 East 26th Street
Commerce, CA 90040
Attn: Hilda Menendez
Phone: 323-724-8223

Customer Code: CHAMELEON
Folder #: 359952
Project: ANNUAL
Sample Group: Annual Screen
Project Manager: Linda Geddes
Phone: (626) 386-1163

The following samples were received from you on **March 29, 2011**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using MWH Laboratories.

Sample #	Sample ID	Sample Date
	@DBP_28 -- Disinfection ByProducts by 300.0	
	@HAA -- Haloacetic Acids	
	@ICP -- ICP Metals	
	@ICPMS -- ICPMS Metals	
	@R226EDD -- Radium 226 (Sub)	
	@R228EDD -- Radium 228 (Sub)	
	@VOA -- Volatile Organics by GCMS	
	@YMOLD -- Yeast and Mold	



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_CA Report
359952

Client: Chameleon Beverage Company, Inc.
6444 East 26th Street
Commerce, CA 90040

Report Date: 05/17/2011
Date Received: 03/29/2011
Sample No: 201103290086

Attention: Hilda Menendez

Sample Id: Spring Water CBC 11 084 0127 EXP MAR 25 13
Date Sampled: 3/29/2011

INVESTIGATION: Analysis per Title 21, Federal Code of Regulations 103.35 - California Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Result	SOQ
201103290086				
GROUP I				
PHYSICAL				
Apparent Color	SM 2120B	3	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1	1.0 TON	3
pH	SM 4500 HB	0.1	6.4	---
Total Dissolved Solid (TDS)	SM 2540C	10	71	500.0
Turbidity	EPA 180.1	0.05	ND NTU	5
GROUP II				
CHEMICAL SUBSTANCE 1				
			Milligrams per Liter	
Alkalinity in CaCO3 units	SM 2320B	2	26	---
Aluminum Total ICAP/MS	EPA 200.8	0.02	ND	0.2
Antimony Total ICAP/MS	EPA 200.8	0.001	ND	0.006
Arsenic Total ICAP/MS	EPA 200.8	0.001	ND	0.010
Asbestos by TEM - >10 microns	EPA 100.2	0.2	ND	---
Barium Total ICAP/MS	EPA 200.8	0.002	0.014	2
Beryllium Total ICAP/MS	EPA 200.8	0.001	ND	0.004
Cadmium Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Chloride	EPA 300.0	1	4.5	250
Chromium Total ICAP/MS	EPA 200.8	0.001	ND	0.1
Copper Total ICAP/MS	EPA 200.8	0.002	ND	1.0
Cyanide	SM 4500CN-F	0.03	ND	0.2
Cyanide by manual distillation	EPA 335.4	0.005	ND	0.2
Fluoride	SM 4500F-C	0.05	ND	1.4-2.4
Iron Total ICAP	EPA 200.7	0.02	ND	0.3
Corrosivity	SM 2330B	-14	-2.800000	---

6/37

SOQ=Standard of Quality per FDA or California
ND=Not detected at the specified limit
All results reported in milligrams per liter unless otherwise noted

Parameter	Method	Reporting Limit	Result	SOQ
Lead Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Magnesium Total ICAP	EPA 200.7	0.0001	1.7	---
Manganese Total ICAP/MS	EPA 200.8	0.002	ND	0.05
Mercury	EPA 245.1	0.0002	ND	0.002
Nickel Total ICAP/MS	EPA 200.8	0.005	ND	0.1
Nitrate-N	EPA 300.0	0.1	0.55	10
Nitrite-N	EPA 300.0	0.05	ND	1
Phenol	EPA 420.4	0.001	ND	0.001
Selenium Total ICAP/MS	EPA 200.8	0.005	ND	0.05
Silver Total ICAP/MS	EPA 200.8	0.0005	ND	0.1
Sodium Total ICAP	EPA 200.7	1	7.5	---
Sulfate	EPA 300.0	0.5	3.4	250.0
Surfactants	SM 5540C	0.05	ND	---
Thallium Total ICAP/MS	EPA 200.8	0.001	ND	0.002
Total Hardness as CaCO3	EPA 200.7	3	19	---
Total Nitrate + Nitrite	EPA 300.0	0.1	0.55	10
Zinc Total ICAP/MS	EPA 200.8	0.02	ND	5.0
GROUP III			Milligrams per Liter	
CHEMICAL SUBSTANCE 2 (VOC)				
1,1,1,2-Tetrachloroethane	EPA 524.2	0.0005	ND	---
1,1,1-Trichloroethane	EPA 524.2	0.0005	ND	0.20
1,1,2,2-Tetrachloroethane	EPA 524.2	0.0005	ND	---
1,1,2-Trichloroethane	EPA 524.2	0.0005	ND	0.005
1,1-Dichloroethane	EPA 524.2	0.0005	ND	---
1,1-Dichloroethene	EPA 524.2	0.0005	ND	0.007
1,1-Dichloropropene	EPA 524.2	0.0005	ND	---
1,2,3-Trichlorobenzene	EPA 524.2	0.0005	ND	---
1,2,3-Trichloropropane	EPA 524.2	0.0005	ND	---
1,2,4-Trichlorobenzene	EPA 524.2	0.0005	ND	0.07
1,2,4-Trimethylbenzene	EPA 524.2	0.0005	ND	---
1,2-Dichloroethane	EPA 524.2	0.0005	ND	0.005
1,2-Dichloropropane	EPA 524.2	0.0005	ND	0.005
1,3,5-Trimethylbenzene	EPA 524.2	0.0005	ND	---
1,3-Dichloropropane	EPA 524.2	0.0005	ND	---
2,2-Dichloropropane	EPA 524.2	0.0005	ND	---
Benzene	EPA 524.2	0.0005	ND	0.005
Bromobenzene	EPA 524.2	0.0005	ND	---
Bromochloromethane	EPA 524.2	0.0005	ND	---
Bromodichloromethane	EPA 524.2	0.0005	ND	---
Bromoform	EPA 524.2	0.0005	ND	---

Parameter	Method	Reporting Limit	Result	SOQ
Bromomethane	EPA 524.2	0.0005	ND	---
Carbon Tetrachloride	EPA 524.2	0.0005	ND	0.005
Chlorobenzene	EPA 524.2	0.0005	ND	0.1
Chlorodibromomethane	EPA 524.2	0.0005	ND	---
Chloroethane	EPA 524.2	0.0005	ND	---
Chloroform (Trichloromethane)	EPA 524.2	0.0005	ND	---
Chloromethane	EPA 524.2	0.0005	ND	---
cis-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.07
cis-1,3-Dichloropropene	EPA 524.2	0.0005	ND	---
Dibromomethane	EPA 524.2	0.0005	ND	---
Dichlorodifluoromethane	EPA 524.2	0.0005	ND	---
Dichloromethane	EPA 524.2	0.0005	ND	0.005
Ethyl benzene	EPA 524.2	0.0005	ND	0.7
Hexachlorobutadiene	EPA 524.2	0.0005	ND	---
Isopropylbenzene	EPA 524.2	0.0005	ND	---
m,p-Xylenes	EPA 524.2	0.0005	ND	---
m-Dichlorobenzene (1,3-DCB)	EPA 524.2	0.0005	ND	---
MTBE	EPA 524.2	0.0005	ND	---
n-Butylbenzene	EPA 524.2	0.0005	ND	---
n-Propylbenzene	EPA 524.2	0.0005	ND	---
o-Chlorotoluene	EPA 524.2	0.0005	ND	---
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.0005	ND	0.6
o-Xylene	EPA 524.2	0.0005	ND	---
p-Chlorotoluene	EPA 524.2	0.0005	ND	---
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.0005	ND	0.075
p-Isopropyltoluene	EPA 524.2	0.0005	ND	---
sec-Butylbenzene	EPA 524.2	0.0005	ND	---
Styrene	EPA 524.2	0.0005	ND	0.1
tert-Butylbenzene	EPA 524.2	0.0005	ND	---
Tetrachloroethylene (PCE)	EPA 524.2	0.0005	ND	0.005
Toluene	EPA 524.2	0.0005	ND	1
Total 1,3-Dichloropropene	EPA 524.2	0.0005	ND	0.0005
Total THM	EPA 524.2	0.0005	ND	0.010
Total xylenes	EPA 524.2	0.001	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.1
trans-1,3-Dichloropropene	EPA 524.2	0.0005	ND	---
Trichloroethylene (TCE)	EPA 524.2	0.0005	ND	0.005
Trichlorotrifluoroethane(Freon)	EPA 524.2	0.0005	ND	---
Vinyl chloride (VC)	EPA 524.2	0.0003	ND	0.002

Parameter	Method	Reporting Limit	Result	SOQ
GROUP IV CHEMICAL SUBSTANCE 3 (NON VOC)			Milligrams per Liter	
2,3,7,8-TCDD, ug/L	EPA 1613B	0.000005	ND	0.000003
2,4,5-TP (Silvex)	EPA 515.4	0.0002	ND	0.05
2,4-D	EPA 515.4	0.0001	ND	0.07
3-Hydroxycarbofuran	EPA 531.2	0.0005	ND	---
Alachlor (Alanex)	EPA 505	0.0001	ND	0.002
Aldicarb (Temik)	EPA 531.2	0.0005	ND	---
Aldicarb sulfone	EPA 531.2	0.0005	ND	---
Aldicarb sulfoxide	EPA 531.2	0.0005	ND	---
Atrazine	EPA 525.2	0.00005	ND	0.003
Baygon	EPA 531.2	0.0005	ND	---
Bentazon	EPA 515.4	0.0005	ND	0.018
Benzo(a)pyrene	EPA 525.2	0.00002	ND	0.0002
Butachlor	EPA 525.2	0.00005	ND	---
Carbaryl	EPA 531.2	0.0005	ND	---
Carbofuran	EPA 531.2	0.0005	ND	40
Chlordane	EPA 505	0.0001	ND	0.002
Dalapon	EPA 515.4	0.001	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.0006	ND	0.4
Di(2-Ethylhexyl)phthalate	EPA 525.2	0.0006	ND	0.004
Dibromochloropropane (DBCP)	EPA 551.1	0.00001	ND	0.0002
Dinoseb	EPA 515.4	0.0002	ND	0.007
Diquat	EPA 549.2	0.0004	ND	0.02
Endothall	EPA 548.1	0.005	ND	0.1
Endrin	EPA 505	0.00001	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.00001	ND	0.000005
Glyphosate	EPA 547	0.006	ND	0.7
Heptachlor	EPA 505	0.00001	ND	0.0004
Heptachlor Epoxide	EPA 505	0.00001	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.00005	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.00005	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.00001	ND	0.0002
Methiocarb	EPA 531.2	0.0005	ND	---
Methomyl	EPA 531.2	0.0005	ND	---
Methoxychlor	EPA 505	0.00005	ND	0.04
Metolachlor	EPA 525.2	0.00005	ND	---
Metribuzin	EPA 525.2	0.00005	ND	---
Oxamyl (Vydate)	EPA 531.2	0.0005	ND	0.2
Pentachlorophenol	EPA 515.4	0.00004	ND	0.001

Picloram	EPA 515.4	0.0001	ND	0.5
Simazine	EPA 525.2	0.00005	ND	0.004
Thiobencarb	EPA 525.2	0.0002	ND	---
Total PCBs	EPA 505	0.0001	ND	0.5
Toxaphene	EPA 505	0.0005	ND	0.003

GROUP V
RADIOACTIVITY
Picocuries per Liter

Alpha, Gross	EPA 900.0	3.0	ND	15
Beta, Gross	EPA 900.0	3.0	ND	50
Radium 226 (Subbed)	EPA 903.1	0.96	<0.960	5
Radium 228 (Subbed)	EPA 904.0	0.80	0.87	5
Uranium ICAP/MS (mg/L)	EPA 200.8	0.001	ND	0.03

GROUP VI
BACTERIOLOGICAL
Milligrams per Liter

E. Coli Bacteria	SM 9223	1.1	<1.1	1.1
Heterotrophic Plate Count	SM 9215B	1	<1	---
Total Coliform Bacteria	SM 9223	1.1	<1.1	2.2

GROUP VII
Disinfection Byproducts
Milligrams per Liter

Bromate by UV/VIS	EPA 317	0.001	ND	0.01
Bromide	EPA 300.0	0.005	0.016	---
Chlorite	EPA 300.0	0.01	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.002	ND	0.06

GROUP VIII
Residual Disinfectants
Milligrams per Liter

Chloramines	SM 4500CL-G/HACH	0.1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.2	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.1	ND	4

Client: Chameleon Beverage Company, Inc.
6444 East 26th Street
Commerce, CA 90040

REPORT DATE: 05/17/2011
DATE RECEIVED: 03/29/2011
SAMPLE NO: 201103290086

Attention: Hilda Menendez

Sample Id: Spring Water CBC 11 084 0127 EXP MAR 25 13

Date Sampled: 3/29/2011

INVESTIGATION: Analysis per Title 21, Federal Code of Regulations 103.35 - New York Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Result	SOQ
GROUP I				
PHYSICAL				
Apparent Color	SM 2120B	3	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1	1.0 TON	3
pH	SM 4500 HB	0.1	6.4	---
Total Dissolved Solid (TDS)	SM 2540C	10	71.0	500
Turbidity	EPA 180.1	0.05	ND NTU	5
GROUP II				
CHEMICAL SUBSTANCE 1				
			Milligrams per Liter	
Alkalinity in CaCO3 units	SM 2320B	2	26.0	---
Aluminum Total ICAP/MS	EPA 200.8	0.02	ND	0.2
Antimony Total ICAP/MS	EPA 200.8	0.001	ND	0.006
Arsenic Total ICAP/MS	EPA 200.8	0.001	ND	0.01
Asbestos by TEM - >10 microns	EPA 100.2	0.2	ND	---
Barium Total ICAP/MS	EPA 200.8	0.002	0.014	2
Beryllium Total ICAP/MS	EPA 200.8	0.001	ND	0.004
Cadmium Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Chloride	EPA 300.0	1	4.5	250
Chromium Total ICAP/MS	EPA 200.8	0.001	ND	0.1
Copper Total ICAP/MS	EPA 200.8	0.002	ND	1
Cyanide	SM 4500CN-F	0.03	ND	0.2
Cyanide by manual distillation	EPA 335.4	0.005	ND	0.2
Fluoride	SM 4500F-C	0.05	ND	1.4-2.4
Iron Total ICAP	EPA 200.7	0.02	ND	0.3
Corrosivity	SM 2330B	-14	-2.8	---
Lead Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Magnesium Total ICAP	EPA 200.7	0.0001	1.7	---
Manganese Total ICAP/MS	EPA 200.8	0.002	ND	0.05
Mercury	EPA 245.1	0.0002	ND	0.002

Parameter	Method	Reporting Limit	Result	SOQ
Nickel Total ICAP/MS	EPA 200.8	0.005	ND	0.1
Nitrate-N	EPA 300.0	0.1	0.55	10
Nitrite-N	EPA 300.0	0.05	ND	1
Phenol	EPA 420.4	0.001	ND	0.001
Selenium Total ICAP/MS	EPA 200.8	0.005	ND	0.05
Silver Total ICAP/MS	EPA 200.8	0.0005	ND	0.1
Sodium Total ICAP	EPA 200.7	1	7.5	---
Sulfate	EPA 300.0	0.5	3.4	250
Surfactants	SM 5540C	0.05	ND	---
Thallium Total ICAP/MS	EPA 200.8	0.001	ND	0.002
Total Hardness as CaCO3	EPA 200.7	3	19.0	---
Total Nitrate + Nitrite	EPA 300.0	0.1	0.55	10
Zinc Total ICAP/MS	EPA 200.8	0.02	ND	5
GROUP III CHEMICAL SUBSTANCE 2 (VOC)			Milligrams per Liter	
1,1,1,2-Tetrachloroethane	EPA 524.2	0.0005	ND	---
1,1,1-Trichloroethane	EPA 524.2	0.0005	ND	0.2
1,1,2,2-Tetrachloroethane	EPA 524.2	0.0005	ND	---
1,1,2-Trichloroethane	EPA 524.2	0.0005	ND	0.005
1,1-Dichloroethane	EPA 524.2	0.0005	ND	---
1,1-Dichloroethene	EPA 524.2	0.0005	ND	0.007
1,1-Dichloropropene	EPA 524.2	0.0005	ND	---
1,2,3-Trichlorobenzene	EPA 524.2	0.0005	ND	---
1,2,3-Trichloropropane	EPA 524.2	0.0005	ND	---
1,2,4-Trichlorobenzene	EPA 524.2	0.0005	ND	0.07
1,2,4-Trimethylbenzene	EPA 524.2	0.0005	ND	---
1,2-Dichloroethane	EPA 524.2	0.0005	ND	0.005
1,2-Dichloropropane	EPA 524.2	0.0005	ND	---
1,3,5-Trimethylbenzene	EPA 524.2	0.0005	ND	---
1,3-Dichloropropane	EPA 524.2	0.0005	ND	---
2,2-Dichloropropane	EPA 524.2	0.0005	ND	---
Benzene	EPA 524.2	0.0005	ND	0.005
Bromobenzene	EPA 524.2	0.0005	ND	---
Bromochloromethane	EPA 524.2	0.0005	ND	---
Bromodichloromethane	EPA 524.2	0.0005	ND	---
Bromoform	EPA 524.2	0.0005	ND	---
Bromomethane	EPA 524.2	0.0005	ND	---
Carbon Tetrachloride	EPA 524.2	0.0005	ND	0.005
Chlorobenzene	EPA 524.2	0.0005	ND	0.1
Chlorodibromomethane	EPA 524.2	0.0005	ND	---
Chloroethane	EPA 524.2	0.0005	ND	---

Parameter	Method	Reporting Limit	Result	SOQ
Chloroform (Trichloromethane)	EPA 524.2	0.0005	ND	---
Chloromethane	EPA 524.2	0.0005	ND	---
cis-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.07
cis-1,3-Dichloropropene	EPA 524.2	0.0005	ND	---
Dibromomethane	EPA 524.2	0.0005	ND	---
Dichlorodifluoromethane	EPA 524.2	0.0005	ND	---
Dichloromethane	EPA 524.2	0.0005	ND	0.005
Ethyl benzene	EPA 524.2	0.0005	ND	0.7
Hexachlorobutadiene	EPA 524.2	0.0005	ND	---
Isopropylbenzene	EPA 524.2	0.0005	ND	---
m,p-Xylenes	EPA 524.2	0.0005	ND	---
m-Dichlorobenzene (1,3-DCB)	EPA 524.2	0.0005	ND	---
MTBE	EPA 524.2	0.0005	ND	---
n-Butylbenzene	EPA 524.2	0.0005	ND	---
n-Propylbenzene	EPA 524.2	0.0005	ND	---
o-Chlorotoluene	EPA 524.2	0.0005	ND	---
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.0005	ND	0.6
o-Xylene	EPA 524.2	0.0005	ND	---
p-Chlorotoluene	EPA 524.2	0.0005	ND	---
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.0005	ND	0.075
p-Isopropyltoluene	EPA 524.2	0.0005	ND	---
sec-Butylbenzene	EPA 524.2	0.0005	ND	---
Styrene	EPA 524.2	0.0005	ND	0.1
tert-Butylbenzene	EPA 524.2	0.0005	ND	---
Tetrachloroethylene (PCE)	EPA 524.2	0.0005	ND	0.005
Toluene	EPA 524.2	0.0005	ND	1
Total 1,3-Dichloropropene	EPA 524.2	0.0005	ND	---
Total THM	EPA 524.2	0.0005	ND	0.08
Total xylenes	EPA 524.2	0.001	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.1
trans-1,3-Dichloropropene	EPA 524.2	0.0005	ND	---
Trichloroethylene (TCE)	EPA 524.2	0.0005	ND	0.005
Trichlorotrifluoroethane(Freon	EPA 524.2	0.0005	ND	---
Vinyl chloride (VC)	EPA 524.2	0.0003	ND	0.002
GROUP IV				
CHEMICAL SUBSTANCE 3 (NON VOC)			Milligrams per Liter	
2,3,7,8-TCDD, ug/L	EPA 1613B	0.000000005	ND	3e-005
2,4,5-TP (Silvex)	EPA 515.4	0.0002	ND	0.05
2,4-D	EPA 515.4	0.0001	ND	0.07
3-Hydroxycarbofuran	EPA 531.2	0.0005	ND	---
Alachlor (Alanex)	EPA 505	0.0001	ND	0.002

Parameter	Method	Reporting Limit	Result	SOQ
Aldicarb (Temik)	EPA 531.2	0.0005	ND	---
Aldicarb sulfone	EPA 531.2	0.0005	ND	---
Aldicarb sulfoxide	EPA 531.2	0.0005	ND	---
Atrazine	EPA 525.2	0.00005	ND	0.003
Baygon	EPA 531.2	0.0005	ND	---
Bentazon	EPA 515.4	0.0005	ND	---
Benzo(a)pyrene	EPA 525.2	0.00002	ND	0.0002
Butachlor	EPA 525.2	0.00005	ND	---
Carbaryl	EPA 531.2	0.0005	ND	---
Carbofuran	EPA 531.2	0.0005	ND	0.04
Chlordane	EPA 505	0.0001	ND	0.002
Dalapon	EPA 515.4	0.001	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.0006	ND	0.4
Di(2-Ethylhexyl)phthalate	EPA 525.2	0.0006	ND	---
Dibromochloropropane (DBCP)	EPA 551.1	0.00001	ND	0.0002
Dinoseb	EPA 515.4	0.0002	ND	0.007
Diquat	EPA 549.2	0.0004	ND	0.02
Endothall	EPA 548.1	0.005	ND	0.1
Endrin	EPA 505	0.00001	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.00001	ND	5e-005
Glyphosate	EPA 547	0.006	ND	0.7
Heptachlor	EPA 505	0.00001	ND	0.0004
Heptachlor Epoxide	EPA 505	0.00001	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.00005	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.00005	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.00001	ND	0.0002
Methiocarb	EPA 531.2	0.0005	ND	---
Methomyl	EPA 531.2	0.0005	ND	---
Methoxychlor	EPA 505	0.00005	ND	0.04
Metolachlor	EPA 525.2	0.00005	ND	---
Metribuzin	EPA 525.2	0.00005	ND	---
Oxamyl (Vydate)	EPA 531.2	0.0005	ND	0.2
Pentachlorophenol	EPA 515.4	0.00004	ND	0.001
Picloram	EPA 515.4	0.0001	ND	0.0005
Simazine	EPA 525.2	0.00005	ND	0.004
Thiobencarb	EPA 525.2	0.0002	ND	---
Total PCBs	EPA 505	0.0001	ND	0.0005
Toxaphene	EPA 505	0.0005	ND	0.003
GROUP V RADIOACTIVITY			Picocuries	
Alpha, Gross	EPA 900.0	3.0	ND	15

SOQ=Standard of Quality per FDA or New York
 ND=Not detected at the specified limit
 All results reported in milligrams per liter unless otherwise noted

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_NY Report
359952

REPORT DATE: 05/17/2011
DATE RECEIVED: 03/29/2011
LABORATORY NO: 201103290086

Beta, Gross	EPA 900.0	3.0	ND	50
Radium 226 (Subbed)	EPA 903.1	0.96	<0.960	3
Radium 228 (Subbed)	EPA 904.0	0.80	0.87	5
Uranium ICAP/MS (mg/L)	EPA 200.8	0.001	ND	0.03
GROUP VI				
BACTERIOLOGICAL				
E. Coli Bacteria	SM 9223	1.1	<1.1	1.1
Heterotrophic Plate Count	SM 9215B	1	<1	---
Total Coliform Bacteria	SM 9223	1.1	<1.1	2.2
GROUP VII				
Disinfection Byproducts				
Bromate by UV/VIS	EPA 317	0.001	ND	0.01
Bromide	EPA 300.0	0.005	0.016	---
Chlorite	EPA 300.0	0.01	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.002	ND	0.06
GROUP VIII				
Residual Disinfectants				
Chloramines	SM 4500CL-G/HACH	0.1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.2	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.1	ND	4

Milligrams per Liter

Milligrams per Liter

Milligrams per Liter



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_CA Report
359952

Client: Chameleon Beverage Company, Inc.
6444 East 26th Street
Commerce, CA 90040

Report Date: 05/17/2011
Date Received: 03/29/2011
Sample No: 201103290086

Attention: Hilda Menendez

Sample Id: Spring Water CBC 11 084 0127 EXP MAR 25 13
Date Sampled: 3/29/2011

INVESTIGATION: Analysis per Title 21, Federal Code of Regulations 103.35 - California Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Result	SOQ
201103290086				
GROUP I				
PHYSICAL				
Apparent Color	SM 2120B	3	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1	1.0 TON	3
Total Dissolved Solid (TDS)	SM 2540C	10	71	500.0
Turbidity	EPA 180.1	0.05	ND NTU	5
GROUP II				
CHEMICAL SUBSTANCE 1				
			Milligrams per Liter	
Aluminum Total ICAP/MS	EPA 200.8	0.02	ND	0.2
Antimony Total ICAP/MS	EPA 200.8	0.001	ND	0.006
Arsenic Total ICAP/MS	EPA 200.8	0.001	ND	0.010
Barium Total ICAP/MS	EPA 200.8	0.002	0.014	2
Beryllium Total ICAP/MS	EPA 200.8	0.001	ND	0.004
Cadmium Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Chloride	EPA 300.0	1	4.5	250
Chromium Total ICAP/MS	EPA 200.8	0.001	ND	0.1
Copper Total ICAP/MS	EPA 200.8	0.002	ND	1.0
Cyanide	SM 4500CN-F	0.03	ND	0.2
Cyanide by manual distillation	EPA 335.4	0.005	ND	0.2
Fluoride	SM 4500F-C	0.05	ND	1.4-2.4
Iron Total ICAP	EPA 200.7	0.02	ND	0.3
Lead Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Manganese Total ICAP/MS	EPA 200.8	0.002	ND	0.05
Mercury	EPA 245.1	0.0002	ND	0.002

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_CA Report
359952

Report Date: 05/17/2011
Date Received: 03/29/2011
Sample No: 201103290086

Parameter	Method	Reporting Limit	Result	SOQ
Nickel Total ICAP/MS	EPA 200.8	0.005	ND	0.1
Nitrate-N	EPA 300.0	0.1	0.55	10
Nitrite-N	EPA 300.0	0.05	ND	1
Phenol	EPA 420.4	0.001	ND	0.001
Selenium Total ICAP/MS	EPA 200.8	0.005	ND	0.05
Silver Total ICAP/MS	EPA 200.8	0.0005	ND	0.1
Sulfate	EPA 300.0	0.5	3.4	250.0
Thallium Total ICAP/MS	EPA 200.8	0.001	ND	0.002
Total Nitrate + Nitrite	EPA 300.0	0.1	0.55	10
Zinc Total ICAP/MS	EPA 200.8	0.02	ND	5.0
GROUP III			Milligrams per Liter	
CHEMICAL SUBSTANCE 2 (VOC)				
1,1,1-Trichloroethane	EPA 524.2	0.0005	ND	0.20
1,1,2-Trichloroethane	EPA 524.2	0.0005	ND	0.005
1,1-Dichloroethene	EPA 524.2	0.0005	ND	0.007
1,2,4-Trichlorobenzene	EPA 524.2	0.0005	ND	0.07
1,2-Dichloroethane	EPA 524.2	0.0005	ND	0.005
1,2-Dichloropropane	EPA 524.2	0.0005	ND	0.005
Benzene	EPA 524.2	0.0005	ND	0.005
Carbon Tetrachloride	EPA 524.2	0.0005	ND	0.005
Chlorobenzene	EPA 524.2	0.0005	ND	0.1
cis-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.07
Dichloromethane	EPA 524.2	0.0005	ND	0.005
Ethyl benzene	EPA 524.2	0.0005	ND	0.7
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.0005	ND	0.6
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.0005	ND	0.075
Styrene	EPA 524.2	0.0005	ND	0.1
Tetrachloroethylene (PCE)	EPA 524.2	0.0005	ND	0.005
Toluene	EPA 524.2	0.0005	ND	1
Total 1,3-Dichloropropene	EPA 524.2	0.0005	ND	0.0005
Total THM	EPA 524.2	0.0005	ND	0.010
Total xylenes	EPA 524.2	0.001	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.1
Trichloroethylene (TCE)	EPA 524.2	0.0005	ND	0.005
Vinyl chloride (VC)	EPA 524.2	0.0003	ND	0.002
GROUP IV			Milligrams per Liter	
CHEMICAL SUBSTANCE 3 (NON VOC)				
2,3,7,8-TCDD, ug/L	EPA 1613B	0.000005	ND	0.000003

SOQ=Standard of Quality per FDA or California
ND=Not detected at the specified limit
All results reported in milligrams per liter unless otherwise noted



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_CA Report
359952

Report Date: 05/17/2011
Date Received: 03/29/2011
Sample No: 201103290086

Parameter	Method	Reporting Limit	Result	SOQ
2,4,5-TP (Silvex)	EPA 515.4	0.0002	ND	0.05
2,4-D	EPA 515.4	0.0001	ND	0.07
Alachlor (Alanex)	EPA 505	0.0001	ND	0.002
Atrazine	EPA 525.2	0.00005	ND	0.003
Bentazon	EPA 515.4	0.0005	ND	0.018
Benzo(a)pyrene	EPA 525.2	0.00002	ND	0.0002
Carbofuran	EPA 531.2	0.0005	ND	40
Chlordane	EPA 505	0.0001	ND	0.002
Dalapon	EPA 515.4	0.001	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.0006	ND	0.4
Di(2-Ethylhexyl)phthalate	EPA 525.2	0.0006	ND	0.004
Dibromochloropropane (DBCP)	EPA 551.1	0.00001	ND	0.0002
Dinoseb	EPA 515.4	0.0002	ND	0.007
Diquat	EPA 549.2	0.0004	ND	0.02
Endothall	EPA 548.1	0.005	ND	0.1
Endrin	EPA 505	0.00001	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.00001	ND	0.000005
Glyphosate	EPA 547	0.006	ND	0.7
Heptachlor	EPA 505	0.00001	ND	0.0004
Heptachlor Epoxide	EPA 505	0.00001	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.00005	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.00005	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.00001	ND	0.0002
Methoxychlor	EPA 505	0.00005	ND	0.04
Oxamyl (Vydate)	EPA 531.2	0.0005	ND	0.2
Pentachlorophenol	EPA 515.4	0.00004	ND	0.001
Picloram	EPA 515.4	0.0001	ND	0.5
Simazine	EPA 525.2	0.00005	ND	0.004
Total PCBs	EPA 505	0.0001	ND	0.5
Toxaphene	EPA 505	0.0005	ND	0.003
GROUP V RADIOACTIVITY			Picocuries per Liter	
Alpha, Gross	EPA 900.0	3.0	ND	15
Beta, Gross	EPA 900.0	3.0	ND	50
Radium 226 (Subbed)	EPA 903.1	0.96	<0.960	5
Radium 228 (Subbed)	EPA 904.0	0.80	0.87	5
Uranium ICAP/MS (mg/L)	EPA 200.8	0.001	ND	0.03

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_CA Report
359952

Report Date: 05/17/2011
Date Received: 03/29/2011
Sample No: 201103290086

GROUP VI			Milligrams per Liter	
BACTERIOLOGICAL				
E. Coli Bacteria	SM 9223	1.1	<1.1	1.1
Total Coliform Bacteria	SM 9223	1.1	<1.1	2.2
GROUP VII			Milligrams per Liter	
Disinfection Byproducts				
Bromate by UV/VIS	EPA 317	0.001	ND	0.01
Chlorite	EPA 300.0	0.01	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.002	ND	0.06
GROUP VIII			Milligrams per Liter	
Residual Disinfectants				
Chloramines	SM 4500CL-G/HACH	0.1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.2	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.1	ND	4

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Client: Chameleon Beverage Company, Inc.
6444 East 26th Street
Commerce, CA 90040

REPORT DATE: 05/17/2011
DATE RECEIVED: 03/29/2011
SAMPLE NO: 201103290086

Attention: Hilda Menendez
Sample Id: Spring Water CBC 11 084 0127 EXP MAR 25 13
Date Sampled: 3/29/2011

INVESTIGATION: Analysis per Title 21, Federal Code of Regulations 103.35 - New York Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Result	SOQ
GROUP I				
PHYSICAL				
Apparent Color	SM 2120B	3	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1	1.0 TON	3
Total Dissolved Solid (TDS)	SM 2540C	10	71.0	500
Turbidity	EPA 180.1	0.05	ND NTU	5
GROUP II				
CHEMICAL SUBSTANCE 1				
			Milligrams per Liter	
Aluminum Total ICAP/MS	EPA 200.8	0.02	ND	0.2
Antimony Total ICAP/MS	EPA 200.8	0.001	ND	0.006
Arsenic Total ICAP/MS	EPA 200.8	0.001	ND	0.01
Barium Total ICAP/MS	EPA 200.8	0.002	0.014	2
Beryllium Total ICAP/MS	EPA 200.8	0.001	ND	0.004
Cadmium Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Chloride	EPA 300.0	1	4.5	250
Chromium Total ICAP/MS	EPA 200.8	0.001	ND	0.1
Copper Total ICAP/MS	EPA 200.8	0.002	ND	1
Cyanide	SM 4500CN-F	0.03	ND	0.2
Cyanide by manual distillation	EPA 335.4	0.005	ND	0.2
Fluoride	SM 4500F-C	0.05	ND	1.4-2.4
Iron Total ICAP	EPA 200.7	0.02	ND	0.3
Lead Total ICAP/MS	EPA 200.8	0.0005	ND	0.005
Manganese Total ICAP/MS	EPA 200.8	0.002	ND	0.05
Mercury	EPA 245.1	0.0002	ND	0.002
Nickel Total ICAP/MS	EPA 200.8	0.005	ND	0.1
Nitrate-N	EPA 300.0	0.1	0.55	10
Nitrite-N	EPA 300.0	0.05	ND	1
Phenol	EPA 420.4	0.001	ND	0.001
Selenium Total ICAP/MS	EPA 200.8	0.005	ND	0.05

20/37



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_NY Report
359952

REPORT DATE: 05/17/2011
DATE RECEIVED: 03/29/2011
LABORATORY NO: 201103290086

Parameter	Method	Reporting Limit	Result	SOQ
Silver Total ICAP/MS	EPA 200.8	0.0005	ND	0.1
Sulfate	EPA 300.0	0.5	3.4	250
Thallium Total ICAP/MS	EPA 200.8	0.001	ND	0.002
Total Nitrate + Nitrite	EPA 300.0	0.1	0.55	10
Zinc Total ICAP/MS	EPA 200.8	0.02	ND	5
GROUP III CHEMICAL SUBSTANCE 2 (VOC)			Milligrams per Liter	
1,1,1-Trichloroethane	EPA 524.2	0.0005	ND	0.2
1,1,2-Trichloroethane	EPA 524.2	0.0005	ND	0.005
1,1-Dichloroethene	EPA 524.2	0.0005	ND	0.007
1,2,4-Trichlorobenzene	EPA 524.2	0.0005	ND	0.07
1,2-Dichloroethane	EPA 524.2	0.0005	ND	0.005
Benzene	EPA 524.2	0.0005	ND	0.005
Carbon Tetrachloride	EPA 524.2	0.0005	ND	0.005
Chlorobenzene	EPA 524.2	0.0005	ND	0.1
cis-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.07
Dichloromethane	EPA 524.2	0.0005	ND	0.005
Ethyl benzene	EPA 524.2	0.0005	ND	0.7
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.0005	ND	0.6
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.0005	ND	0.075
Styrene	EPA 524.2	0.0005	ND	0.1
Tetrachloroethylene (PCE)	EPA 524.2	0.0005	ND	0.005
Toluene	EPA 524.2	0.0005	ND	1
Total THM	EPA 524.2	0.0005	ND	0.08
Total xylenes	EPA 524.2	0.001	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.0005	ND	0.1
Trichloroethylene (TCE)	EPA 524.2	0.0005	ND	0.005
Vinyl chloride (VC)	EPA 524.2	0.0003	ND	0.002
GROUP IV CHEMICAL SUBSTANCE 3 (NON VOC)			Milligrams per Liter	
2,3,7,8-TCDD, ug/L	EPA 1613B	0.00000005	ND	3e-005
2,4,5-TP (Silvex)	EPA 515.4	0.0002	ND	0.05
2,4-D	EPA 515.4	0.0001	ND	0.07
Alachlor (Alanex)	EPA 505	0.0001	ND	0.002
Atrazine	EPA 525.2	0.00005	ND	0.003
Benzo(a)pyrene	EPA 525.2	0.00002	ND	0.0002
Carbofuran	EPA 531.2	0.0005	ND	0.04
Chlordane	EPA 505	0.0001	ND	0.002
Dalapon	EPA 515.4	0.001	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.0006	ND	0.4
Dibromochloropropane (DBCP)	EPA 551.1	0.00001	ND	0.0002



MWH
LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
FDA_NY Report
359952

REPORT DATE: 05/17/2011
DATE RECEIVED: 03/29/2011
LABORATORY NO: 201103290086

Dinoseb	EPA 515.4	0.0002	ND	0.007
Diquat	EPA 549.2	0.0004	ND	0.02
Endothall	EPA 548.1	0.005	ND	0.1
Endrin	EPA 505	0.00001	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.00001	ND	5e-005
Glyphosate	EPA 547	0.006	ND	0.7
Heptachlor	EPA 505	0.00001	ND	0.0004
Heptachlor Epoxide	EPA 505	0.00001	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.00005	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.00005	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.00001	ND	0.0002
Methoxychlor	EPA 505	0.00005	ND	0.04
Oxamyl (Vydate)	EPA 531.2	0.0005	ND	0.2
Pentachlorophenol	EPA 515.4	0.00004	ND	0.001
Picloram	EPA 515.4	0.0001	ND	0.0005
Simazine	EPA 525.2	0.00005	ND	0.004
Total PCBs	EPA 505	0.0001	ND	0.0005
Toxaphene	EPA 505	0.0005	ND	0.003

GROUP V
RADIOACTIVITY

Alpha, Gross	EPA 900.0	3.0	ND	15
Beta, Gross	EPA 900.0	3.0	ND	50
Radium 226 (Subbed)	EPA 903.1	0.96	<0.960	3
Radium 228 (Subbed)	EPA 904.0	0.80	0.87	5
Uranium ICAP/MS (mg/L)	EPA 200.8	0.001	ND	0.03

Picocuries

GROUP VI
BACTERIOLOGICAL

E. Coli Bacteria	SM 9223	1.1	<1.1	1.1
Total Coliform Bacteria	SM 9223	1.1	<1.1	2.2

Milligrams per Liter

GROUP VII
Disinfection Byproducts

Bromate by UV/VIS	EPA 317	0.001	ND	0.01
Chlorite	EPA 300.0	0.01	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.002	ND	0.06

Milligrams per Liter

GROUP VIII
Residual Disinfectants

Chloramines	SM 4500CL-G/HACH	0.1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.2	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.1	ND	4

Milligrams per Liter



MWH

LABORATORIES

A Division of MWH Americas, Inc.

750 Royal Oak Dr., Suite 100
Monrovia, California, 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Chameleon Beverage Company, Inc.
Hilda Menendez
6444 East 26th Street
Commerce, CA 90040

Laboratory Comments
Report: #359952

Group Comments

Analytical results for Radium 226, and Radium 228 are submitted by Pace Analytical Services, Greensburg, PA

May 11, 2011

Ms. Jaclyn L. Contreras
MWH Americas, Inc.
Royal Oaks Dr.
Suite 100
Monrovia, CA 910163629

RE: Project: PACE PA 359952
Pace Project No.: 3045350

Dear Ms. Contreras:

Enclosed are the analytical results for sample(s) received by the laboratory on April 21, 2011. The results relate only to the samples included in this report. Results reported herein conform to the most current NELAC standards, where applicable, unless otherwise narrated in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Carin Ferris for
Jacquelyn Collins
jacquelyn.collins@pacelabs.com
Project Manager

Enclosures

cc: Mr. Aleksandar D. Tomovich, MWH Americas, Inc.

REPORT OF LABORATORY ANALYSIS

24/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



CERTIFICATIONS

Project: PACE PA 359952
Pace Project No.: 3045350

Pennsylvania Certification IDs

1638 Roseytown Road Suites 2,3&4, Greensburg, PA 15601
Alabama Certification #: 41590
Arizona Certification #: AZ0734
Arkansas Certification
California/NELAC Certification #: 04222CA
Colorado Certification
Connecticut Certification #: PH 0694
Delaware Certification
Florida/NELAC Certification #: E87683
Guam/PADEP Certification
Hawaii/PADEP Certification
Idaho Certification
Illinois/PADEP Certification
Indiana/PADEP Certification
Iowa Certification #: 391
Kansas/NELAC Certification #: E-10358
Kentucky Certification #: 90133
Louisiana/NELAC Certification #: LA080002
Louisiana/NELAC Certification #: 4086
Maine Certification #: PA0091
Maryland Certification #: 308
Massachusetts Certification #: M-PA1457

Michigan/PADEP Certification
Missouri Certification #: 235
Montana Certification #: Cert 0082
Nevada Certification
New Hampshire/NELAC Certification #: 2976
New Jersey/NELAC Certification #: PA 051
New Mexico Certification
New York/NELAC Certification #: 10888
North Carolina Certification #: 42706
Oregon/NELAC Certification #: PA200002
Pennsylvania/NELAC Certification #: 65-00282
Puerto Rico Certification #: PA01457
South Dakota Certification
Tennessee Certification #: TN2867
Texas/NELAC Certification #: T104704188-09 TX
Utah/NELAC Certification #: ANTE
Virgin Island/PADEP Certification
Virginia Certification #: 00112
Washington Certification #: C1941
West Virginia Certification #: 143
Wisconsin/PADEP Certification
Wyoming Certification #: 8TMS-Q

REPORT OF LABORATORY ANALYSIS

25/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



SAMPLE SUMMARY

Project: PACE PA 359952

Pace Project No.: 3045350

Lab ID	Sample ID	Matrix	Date Collected	Date Received
3045350001	201103290086	Drinking Water	03/29/11 11:39	04/21/11 10:00

REPORT OF LABORATORY ANALYSIS

26/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: PACE PA 359952

Pace Project No.: 3045350

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3045350001	201103290086	EPA 903.1	RMD	1	PASI-PA
		EPA 904.0	DJL	1	PASI-PA

REPORT OF LABORATORY ANALYSIS

27/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: PACE PA 359952
Pace Project No.: 3045350

Method: EPA 903.1
Description: 903.1 Radium 226
Client: MWH Laboratories
Date: May 11, 2011

General Information:

1 sample was analyzed for EPA 903.1. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

28/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: PACE PA 359952
Pace Project No.: 3045350

Method: EPA 904.0
Description: 904.0 Radium 228
Client: MWH Laboratories
Date: May 11, 2011

General Information:

1 sample was analyzed for EPA 904.0. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

This data package has been reviewed for quality and completeness and is approved for release.

REPORT OF LABORATORY ANALYSIS

29/37

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: PACE PA 359952

Pace Project No.: 3045350

Sample: 201103290086 **Lab ID: 3045350001** Collected: 03/29/11 11:39 Received: 04/21/11 10:00 Matrix: Drinking Water
PWS: Site ID: Sample Type:

Parameters	Method	Act ± Unc (MDC)	Units	Analyzed	CAS No.	Qual
Radium-226	EPA 903.1	0.000 ± 0.428 (0.960)	pCi/L	05/05/11 13:09	13982-63-3	
Radium-228	EPA 904.0	0.872 ± 0.433 (0.799)	pCi/L	05/09/11 12:35	15262-20-1	

QUALITY CONTROL DATA

Project: PACE PA 359952
Pace Project No.: 3045350

QC Batch: RADC/8141	Analysis Method: EPA 903.1
QC Batch Method: EPA 903.1	Analysis Description: 903.1 Radium-226
Associated Lab Samples: 3045350001	

METHOD BLANK: 291340	Matrix: Water
Associated Lab Samples: 3045350001	

Parameter	Act ± Unc (MDC)	Units	Analyzed	Qualifiers
Radium-226	0.237 ± 0.467 (0.851)	pCi/L	05/05/11 12:58	

QUALITY CONTROL DATA

Project: PACE PA 359952

Pace Project No.: 3045350

QC Batch: RADC/8180

Analysis Method: EPA 904.0

QC Batch Method: EPA 904.0

Analysis Description: 904.0 Radium 228

Associated Lab Samples: 3045350001

METHOD BLANK: 292389

Matrix: Water

Associated Lab Samples: 3045350001

Parameter	Act ± Unc (MDC)	Units	Analyzed	Qualifiers
Radium-228	-0.0855 ± 0.355 (0.834)	pCi/L	05/09/11 12:35	

QUALIFIERS

Project: PACE PA 359952

Pace Project No.: 3045350

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to changes in sample preparation, dilution of the sample aliquot, or moisture content.

ND - Not Detected at or above adjusted reporting limit.

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

S - Surrogate

1,2-Diphenylhydrazine (8270 listed analyte) decomposes to Azobenzene.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Act - Activity

Unc - Uncertainty

(MDC) - Minimum Detectable Concentration

Pace Analytical is NELAP accredited. Contact your Pace PM for the current list of accredited analytes.

LABORATORIES

PASI-PA Pace Analytical Services - Greensburg



MWH Laboratories
 A Division of MWH Americas, Inc.
 750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016-3629
 Ph (626) 386-1100 Fax (626) 386-1095

Ship To

1638 Roseytown Road, Suite 2
 PACE Analytical Services, Inc.

Greensburg, PA 15601

724-850-5600

Fax 724-850-5601

MWH Project # **359952** Report Due: **05/12/2011** Sub PO# **99-10141**

JLS



EPA 903.1 R226EDD

EPA 904.0 R228EDD

Client Sample ID for reference only Analysis Requested
 Purified Water CBC 11 084 0127 EXP I Radium 226 (Sub)
 Purified Water CBC 11 084 0127 EXP I Radium 228 (Sub)

Preserved @ 1130 4/21/11

3045350

Date 4/20/2011

Submittal Form & Purchase Order 99-10141

*REPORTING REQUIREMENTS: Do Not Combine Report with any other samples submitted under different MWH project numbers!
 Report & Invoice must have the MWH Project Number 359952 Sub PO# 99-10141 and Job # 1000014



Report all quality control data according to Method, include dates analyzed, date extracted (if extracted) and Method reference on the report.
 Results must have Complete data & QC with Approval Signature. See reverse side for List of Terms and Conditions

Reports: Jackie Contreras Sub-Contracting Administrator
 EMAIL TO: mwhlabs-subcontractreports@mwhglobal.com
 MWH Laboratories 750 Royal Oaks Dr. Ste. 100, Monrovia, CA 91016
 Phone (626) 386-1165 Fax (626) 386-1122
 Invoices to: MWH LABORATORIES
 Accounts Payable PO BOX 6610, Broomfield, CO 80021

Provide in each Report the Specified State
 Certification # & Exp Date for requested tests
 + matrix.

Samples from the State of CALIFORNIA

Sample Date & Time Matrix PWS Systemcode PWSID
 03/29/11 1139 Water 001
 03/29/11 1139 Water

Relinquished by:  Sample Control Date 4/20 Time 12:00 MUST HAVE NOTIFICATION IF TEMP IS GREATER THAN 6 OR LESS THAN CELSIUS
 Received by:  Date 4/21/11 Time 1:00 An Acknowledgement of Receipt is requested to attn: Christine Lewis

813

Sample Condition Upon Receipt



Client Name: MWH

Project # 3045350

Courier: Fed Ex UPS USPS Client Commercial Pace Other _____

Tracking #: 929428746553

Optional
Proj. Due Date:
Proj. Name:

Custody Seal on Cooler/Box Present: yes no Seals intact: yes no

Packing Material: Bubble Wrap Bubble Bags None Other _____

Thermometer Used 3 5 Type of Ice: Wet Blue None Samples on ice, cooling process has begun

Cooler Temperature NA Biological Tissue is Frozen: Yes No

Date and initials of person examining contents: EEH/1/11

Temp should be above freezing to 6°C

Comments:

Chain of Custody Present:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	1.
Chain of Custody Filled Out:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	2.
Chain of Custody Relinquished:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	3.
Sampler Name & Signature on COC:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	6.
Rush Turn Around Time Requested:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	7.
Sufficient Volume:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	8.
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	9.
-Pace Containers Used:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	
Containers Intact:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	10.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11.
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	12.
-Includes date/time/ID/Analysis Matrix: <u>DW</u>		
All containers needing preservation have been checked.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <u>Preserved w/ bml H2O3 @ 11:30 4/21/11</u>
All containers needing preservation are found to be in compliance with EPA recommendation.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	
exceptions: VOA, coliform, TOC, O&G, WI-DRO (water)	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Initial when completed <u>EEH</u> Lot # of added preservative <u>RAF11-0057-5</u>
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14.
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15.
Trip Blank Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Pace Trip Blank Lot # (if purchased):		

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted: _____ Date/Time: _____

Comments/ Resolution: _____

Project Manager Review:

[Signature]

Date: 4/21/11

Note: Whenever there is a discrepancy affecting North Carolina compliance samples, a copy of this form will be sent to the North Carolina DEHNR Certification Office (i.e. out of hold, incorrect preservative, out of temp, incorrect containers)

Quality Control Sample Performance Assessment



Test: Ra-226
Analyst: RMD
Date: 4/26/2011
Worklist: 8141
Matrix: DW

Analyst Must Manually Enter All Fields Highlighted in Yellow.

Method Blank Assessment	
MB Sample ID	291340
MB concentration:	0.237
MB Counting Uncertainty:	0.465
MB MDC:	0.851
MB Numerical Performance Indicator:	4.22
MB Status vs Numerical Indicator:	N/A
MB Status vs. MDC:	Pass

Laboratory Control Sample Assessment		LCS(Y or N)?	Y
Count Date:	5/5/2011	LCS8141	LCS8141
Spike I.D.:	09-036		5/5/2011
Spike Concentration (pCi/mL):	59.735		09-036
Volume Used (mL):	0.10		59.735
Aliquot Volume (L, g, F):	0.500		0.10
Target Conc. (pCi/L, g, F):	11.947		11.947
Uncertainty (Calculated):	0.287		0.287
Result (pCi/L, g, F):	13.892		11.432
LCS/LCSD Counting Uncertainty (pCi/L, g, F):	2.177		1.844
Numerical Performance Indicator:	1.74		-0.53
Percent Recovery:	116.28%		95.69%
Status vs Numerical Indicator:	N/A		N/A
Status vs Recovery:	Pass		Pass

36/37

Duplicate Sample Assessment	
Sample I.D.:	LCS8141
Duplicate Sample I.D.:	LCS8141
Sample Result (pCi/L, g, F):	13.892
Sample Result Counting Uncertainty (pCi/L, g, F):	2.177
Sample Duplicate Result (pCi/L, g, F):	11.432
Sample Duplicate Counting Uncertainty (pCi/L, g, F):	1.844
Are sample and/or duplicate results below MDC?	NO
Duplicate Numerical Performance Indicator:	1.689
Duplicate RPD:	19.42%
Duplicate Status vs Numerical Indicator:	N/A
Duplicate Status vs RPD:	Pass

Evaluation of duplicate precision is not applicable if either the sample or duplicate results are below the MDC.

Comments:

On 5/11/11

Sample Matrix Spike Control Assessment	
Sample Collection Date:	4/21/2011
Sample I.D.:	3045358001
Sample MS I.D.:	3045358001MS
Spike I.D.:	09-036
MS/MSD Decay Corrected Spike Concentration (pCi/mL):	59.736
Spike Volume Used in MS (mL):	0.20
MS Aliquot (L, g, F):	0.500
MS Target Conc (pCi/L, g, F):	23.894
MSD Aliquot (L, g, F):	
MSD Target Conc. (pCi/L, g, F):	
Spike uncertainty (calculated):	0.573
Sample Result:	0.083
Sample Result Counting Uncertainty (pCi/L, g, F):	0.162
Sample Matrix Spike Result:	24.709
Matrix Spike Result Counting Uncertainty (pCi/L, g, F):	2.715
Sample Matrix Spike Duplicate Result:	
Matrix Spike Duplicate Result Counting Uncertainty (pCi/L, g, F):	
MS Numerical Performance Indicator:	0.516
MSD Numerical Performance Indicator:	
MS Percent Recovery:	103.06%
MSD Percent Recovery:	
MS Status vs Numerical Indicator:	N/A
MSD Status vs Numerical Indicator:	
MS Status vs Recovery:	Pass
MSD Status vs Recovery:	

Matrix Spike/Matrix Spike Duplicate Sample Assessment	
Sample I.D.:	
Sample MS I.D.:	
Sample MSD I.D.:	
Sample Matrix Spike Result:	
Sample Matrix Spike Counting Uncertainty (pCi/L, g, F):	
Sample Matrix Spike Duplicate Result:	
Sample Matrix Spike Duplicate Counting Uncertainty (pCi/L, g, F):	
Duplicate Numerical Performance Indicator:	
MS/MSD Duplicate RPD:	
MS/MSD Duplicate Status vs Numerical Indicator:	
MS/MSD Duplicate Status vs RPD:	

Quality Control Sample Performance Assessment



Analyst Must Manually Enter All Fields Highlighted in Yellow.

Test: Ra-228
 Analyst: DJL
 Date: 4/19/2011
 Worklist: 8180
 Matrix: DW

Method Blank Assessment

MB Sample ID: 292389
 MB concentration: -0.086
 MB Counting Uncertainty: 0.354
 MB MDC: 0.834
 MB Numerical Performance Indicator: 0.47
 MB Status vs Numerical Indicator: N/A
 MB Status vs MDC: Pass

Laboratory Control Sample Assessment

LCSD (Y or N)?	Y
Count Date:	5/9/2011
Spike I.D.:	09-037
Spike Concentration (pCi/ml):	73.709
Volume Used (ml):	0.10
Aliquot Volume (L, g, F):	0.800
Target Conc. (pCi/L, g, F):	9.214
Uncertainty (Calculated):	0.286
Result (pCi/L, g, F):	9.747
LCSD Counting Uncertainty (pCi/L, g, F):	1.237
Numerical Performance Indicator:	0.82
Percent Recovery:	105.79%
Status vs Numerical Indicator:	N/A
Status vs Recovery:	Pass

Duplicate Sample Assessment

Sample I.D.: LCS8180
 Duplicate Sample I.D.: LCS8180
 Sample Result (pCi/L, g, F): 9.747
 Sample Result Counting Uncertainty (pCi/L, g, F): 1.237
 Sample Duplicate Result (pCi/L, g, F): 11.249
 Sample Duplicate Counting Uncertainty (pCi/L, g, F): 1.450
 Are sample and/or duplicate results below MDC? NO
 Duplicate Numerical Performance Indicator: -1.544
 Duplicate RPD: 14.31%
 Duplicate Status vs Numerical Indicator: N/A
 Duplicate Status vs RPD: Pass

Sample Matrix Spike Control Assessment

Sample Collection Date: 4/21/2011
 Sample I.D.: 3045361001
 Sample MS I.D.: 3045361001MS
 Sample MSD I.D.: 09-037
 Spike I.D.: 74.148
 MS/MSD Decay Corrected Spike Concentration (pCi/ml): 0.20
 Spike Volume Used in MS (ml): 0.800
 Spike Volume Used in MSD (ml): 18.537
 MS Aliquot (L, g, F): 0.575
 MS Target Conc. (pCi/L, g, F): 1.043
 MSD Aliquot (L, g, F): 23.217
 MSD Target Conc. (pCi/L, g, F): 1.940
 Spike uncertainty (calculated): 3.414
 Sample Result: 119.62%
 Sample Matrix Spike Result: N/A
 Matrix Spike Result Counting Uncertainty (pCi/L, g, F): Pass
 Sample Matrix Spike Duplicate Result: Pass
 Matrix Spike Duplicate Result Counting Uncertainty (pCi/L, g, F): Pass
 MS Numerical Performance Indicator: Pass
 MSD Numerical Performance Indicator: Pass
 MS Percent Recovery: Pass
 MSD Percent Recovery: Pass
 MS Status vs Numerical Indicator: Pass
 MSD Status vs Numerical Indicator: Pass
 MS Status vs Recovery: Pass
 MSD Status vs Recovery: Pass

Matrix Spike/Matrix Spike Duplicate Sample Assessment

Sample I.D.:
 Sample MS I.D.:
 Sample MSD I.D.:
 Matrix Spike Result:
 Matrix Spike Result Counting Uncertainty (pCi/L, g, F):
 Sample Matrix Spike Duplicate Result:
 Sample Matrix Spike Duplicate Counting Uncertainty (pCi/L, g, F):
 Duplicate Numerical Performance Indicator:
 MS/MSD Duplicate RPD:
 MS/MSD Duplicate Status vs Numerical Indicator:
 MS/MSD Duplicate Status vs RPD:

Evaluation of duplicate precision is not applicable if either are sample or duplicate results are below the MDC.

Comments:

04-5/11