

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

Chameleon Beverage Company
Chameleon Beverage Company
6444 E 26TH ST
COMMERCE, CA 90040-3214
Attention: Araceli Ramirez
Fax: 323-724-9048

Date of Issue
08/16/2022

Rachelle Arada
EUROFINS EATON
ANALYTICAL, LLC

L6NW: Rachelle Arada
Project Manager

Report: 997695
Project: ANNUAL
Group: annual-50 state



Utah ELCP CA00006

* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

* Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.

* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

* Test results relate only to the sample(s) tested.

* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

* This report shall not be reproduced except in full, without the written approval of the laboratory.

* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	NE-OS-21-13
Arkansas	CA00006	Nevada	CA00006
California	2813	New Hampshire *	2959
Colorado	CA00006	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	CA00006
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	21-008R	Ohio - 537.1	87786
Hawaii	CA00006	Oregon *	4034
Idaho	CA00006	Pennsylvania *	68-00565
Illinois	200033	Puerto Rico	CA00006
Indiana	C-CA-01	Rhode Island	LAO00326
Iowa – Asbestos	413	South Carolina	87016
Kansas *	E-10268	South Dakota	CA11320
Kentucky	90107	Tennessee	TN02839
Louisiana *	LA008	Texas *	T104704230-20-18
Maine	CA00006	Utah (Primary AB) *	CA00006
Maryland	224	Vermont	VT0114
Marianas Islands	MP0004	Virginia *	460260
Massachusetts	M-CA006	Washington	C838
Michigan	9906	EPA Region 5	CA00006
Mississippi	CA00006	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2917 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

Test(s)	Method(s)	Potable Water *	Waste Water
Enterococci	Enterolert	x	x
<i>Escherichia coli</i> (Enumeration)	SM 9221 B.1 SM 9221 F	x	
Fecal Coliform (P/A and Enumeration)	SM 9221 C (MTF/EC), SM 9221 E (MTF/EC)	x	x
Fecal Streptococci and Enterococci	SM 9230 B	x	x
Heterotrophic Bacteria	SM 9215 B	x	
Legionella	Legiolert®	x	
<i>Pseudomonas aeruginosa</i>	Idexx Pseudalart	x	
Total Coliform (P/A and Enumeration)	SM 9221A, SM 9221B, SM 9221 C	x	x
Total Coliform, Total Coliform with Chlorine Present	SM 9221 B	x	x
Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure)	SM 9223	x	
Total Microcystins and Nodularins	EPA 546	X	
Yeast and Mold	SM 9610	x	
1,2,3-Trichloropropane (TCP) at 5 PPT	CA SRL 524M-TCP	x	
1,4-Dioxane	EPA 522	x	
2,3,7,8-TCDD	Modified EPA 1613 B	x	
Acrylamide	* LCMS 2440)	x	
Algal Toxins/Microcystin	* LCMS 3570	x	
Alkalinity	SM 2320B	x	x
Ammonia	EPA 350.1, SM 4500-NH3 H		x
Asbestos	EPA 100.2	x	x
Bicarbonate Alkalinity as HCO3	SM 2330 B	x	x
BOD/CBOD	SM 5210 B		x
Bromate	* LCMS- 2447	x	
Carbonate as CO3	SM 2330 B	x	x
Carbonyls	EPA 556	x	x
Chemical Oxygen Demand	EPA 410.4, SM 5220D		x
Chlorinated Acids	EPA 515.4	x	
Chlorine Dioxide	Palin Test Chlordio X Plus, SM 4500-CLO2 D	x	
Chlorine, Free, Combined, Total Residual, Chloramines	SM 4500-CI G	x	
Color	SM2120B	x	
Conductivity	EPA 120.1, SM 2510B	x	x
Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated	SM 2330 B	x	
Cyanide (Amenable)	SM 4500-CN G	x	x
Cyanide (Free)	SM 4500CN F	x	x
Cyanide (Total)	EPA 335.4	x	x
Cyanogen Chloride (Screen)	* 335 Mod (WC-24467)	x	
Diquat and Paraquat	EPA 549.2	x	
DBP and HAA	SM 6251 B	x	
Dissolved Organic Carbon	SM 5310 C	x	
Dissolved Oxygen	SM 4500-O G		x
EDB/DCBP/TCP	EPA 504.1	x	
EDB/DBCP and Disinfection Byproducts	EPA 551.1	x	
EDTA and NTA	* WC-2454	x	
Endothall	EPA 548.1, *(LCMS-2445)	x	
Fluoride	SM 4500F C	x	x
Glyphosate	EPA 547	x	
Glyphosate and AMPA	* LCMS-3618	x	
Gross Alpha and Gross Beta	EPA 900.0	x	x

Test(s)	Method(s)	Potable Water *	Waste Water
Gross Alpha coprecipitation	SM 7110 C	x	x
Hardness	SM 2340 B	x	x
Hexavalent Chromium	EPA 218.6,	x	x
Hexavalent Chromium	EPA 218.7,	x	
Hexavalent Chromium	SM 3500-Cr B		x
Inorganic Anions and DBPs	EPA 300.0	x	x
Norganic Anions and DBPs	EPA 300.1	x	
Kjeldahl Nitrogen	EPA 351.2		x
Metals	EPA 200.7, EPA200.8	x	x
Nitrosamines	EEA-Agilent 521.1 (GCMS-24250)	x	
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x
Odor	SM2150B	x	
Organohalide Pesticides and PCB	EPA 505	x	
Ortho Phosphate	SM 4500P E	x	
Oxyhalides Disinfection Byproducts	EPA 317.0	x	
Perchlorate	EPA 331.0	x	
Perchlorate (Low and High Levels)	EPA 314.0	x	
Perfluorinated Alkyl Acids	EPA 533, EPA 537, EPA 537.1	x	
PPCP and EDC	* LCMS-2443	x	
pH	EPA 150.1 SM 4500-H+ B	x	x
Phenolics – Low Level	*WC 2493 (EPA 420.2 and EPA 420.4 MOD)	x	x
Phenylurea Pesticides/Herbicides	* LCMS-2448	x	
Radium-226, Radium-228	GA Tech (Rad-2374)	x	
Radon-222	SM 7500RN	x	
Residue (Filterable)	SM 2540C	x	x
Residue (Non-Filterable)	SM 2540D		x
Residue (Total)	SM 2540B		x
Residue (Volatile)	EPA 160.4		x
Semi-Volatile Compounds	EPA 525.2	x	
Silica	SM 4500-SiO2 C	x	x
Sulfide	SM 4500-S D		x
Sulfite	SM 4500-SO3 B	x	x
Surfactants	SM 5540C	x	x
Taste and Odor	SM 6040 E	x	
Total Organic Carbon	SM 5310 C	x	x
Total Phenols	EPA 420.1		x
Total Phenols	EPA 420.4	x	x
Triazine Pesticides and their Degradates	* LCMS-3617	x	
Turbidity	EPA 180.1	x	x
Uranium by ICP/MS	EPA 200.8	x	
UV 254 Organic Constituents	SM 5910B	x	
VOCs	EPA 524.2	x	
VOCs	*(GCMS 2412) by EPA 524.2 modified	x	

(*) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+) In-House Method

Acknowledgement of Samples Received

Addr: **Chameleon Beverage Company**
 Chameleon Beverage Company
 6444 E 26TH ST
 COMMERCE, CA 90040-3214
 United States

Attn: Araceli Ramirez
 Phone: 3237248223

Client ID: CHAMELEON
 Folder #: 997695
 Project: ANNUAL
 Sample Group: annual-50 state

Project Manager: Rachelle Arada
 Phone: 626-386-1106

The following samples were received from you on **April 08, 2022 at 00:36**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date																																																						
<u>202204080117</u>	Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL	04/08/2022 1010																																																						
	<table> <tr> <td>@2378-TCDD_Dioxin</td><td>@331-PHG</td><td>@504LOW</td></tr> <tr> <td>@505</td><td>@515.4</td><td>@525</td></tr> <tr> <td>@531</td><td>@549</td><td>@551SODA</td></tr> <tr> <td>@ANIONS28</td><td>@ANIONS48</td><td>@DBP_14</td></tr> <tr> <td>@DBP_28</td><td>@HAA</td><td>@ICP</td></tr> <tr> <td>@ICPMS</td><td>@QUANT2000 18HR</td><td>@RA226 GA</td></tr> <tr> <td>@RA228 GA</td><td>@VOA</td><td>Alkalinity in CaCO3 units</td></tr> <tr> <td>Anion Sum - Calculated</td><td>Apparent Color</td><td>Asbestos by TEM - >10 microns</td></tr> <tr> <td>Bicarb.Alkalinity as HCO3,calc</td><td>Bromate by UV/VIS</td><td>Carbon Dioxide,Free(25C)-Calc.</td></tr> <tr> <td>Carbonate as CO3, Calculated</td><td>Cation Sum - Calculated</td><td>Chloramines</td></tr> <tr> <td>Chlorine Dioxide</td><td>Cyanide</td><td>Endothall</td></tr> <tr> <td>Fluoride</td><td>Free Chlorine Residual</td><td>Glyphosate</td></tr> <tr> <td>Heterotrophic Plate Count</td><td>Hydroxide as OH, Calculated</td><td>Langelier Index - 25 degree</td></tr> <tr> <td>Mercury ICPMS</td><td>Odor at 60 C (TON)</td><td>pH of CaCO3 saturation(25C)</td></tr> <tr> <td>PH, Bottled Water</td><td>Phenolic Compounds-low level</td><td>Specific Conductance</td></tr> <tr> <td>Surfactants</td><td>Total Chlorine Residual</td><td>Total Dissolved Solid (TDS)</td></tr> <tr> <td>Total Hardness as CaCO3 by ICP</td><td>Turbidity</td><td>@RADEDD</td></tr> <tr> <td>@YMOLD</td><td></td><td></td></tr> </table>	@2378-TCDD_Dioxin	@331-PHG	@504LOW	@505	@515.4	@525	@531	@549	@551SODA	@ANIONS28	@ANIONS48	@DBP_14	@DBP_28	@HAA	@ICP	@ICPMS	@QUANT2000 18HR	@RA226 GA	@RA228 GA	@VOA	Alkalinity in CaCO3 units	Anion Sum - Calculated	Apparent Color	Asbestos by TEM - >10 microns	Bicarb.Alkalinity as HCO3,calc	Bromate by UV/VIS	Carbon Dioxide,Free(25C)-Calc.	Carbonate as CO3, Calculated	Cation Sum - Calculated	Chloramines	Chlorine Dioxide	Cyanide	Endothall	Fluoride	Free Chlorine Residual	Glyphosate	Heterotrophic Plate Count	Hydroxide as OH, Calculated	Langelier Index - 25 degree	Mercury ICPMS	Odor at 60 C (TON)	pH of CaCO3 saturation(25C)	PH, Bottled Water	Phenolic Compounds-low level	Specific Conductance	Surfactants	Total Chlorine Residual	Total Dissolved Solid (TDS)	Total Hardness as CaCO3 by ICP	Turbidity	@RADEDD	@YMOLD			
@2378-TCDD_Dioxin	@331-PHG	@504LOW																																																						
@505	@515.4	@525																																																						
@531	@549	@551SODA																																																						
@ANIONS28	@ANIONS48	@DBP_14																																																						
@DBP_28	@HAA	@ICP																																																						
@ICPMS	@QUANT2000 18HR	@RA226 GA																																																						
@RA228 GA	@VOA	Alkalinity in CaCO3 units																																																						
Anion Sum - Calculated	Apparent Color	Asbestos by TEM - >10 microns																																																						
Bicarb.Alkalinity as HCO3,calc	Bromate by UV/VIS	Carbon Dioxide,Free(25C)-Calc.																																																						
Carbonate as CO3, Calculated	Cation Sum - Calculated	Chloramines																																																						
Chlorine Dioxide	Cyanide	Endothall																																																						
Fluoride	Free Chlorine Residual	Glyphosate																																																						
Heterotrophic Plate Count	Hydroxide as OH, Calculated	Langelier Index - 25 degree																																																						
Mercury ICPMS	Odor at 60 C (TON)	pH of CaCO3 saturation(25C)																																																						
PH, Bottled Water	Phenolic Compounds-low level	Specific Conductance																																																						
Surfactants	Total Chlorine Residual	Total Dissolved Solid (TDS)																																																						
Total Hardness as CaCO3 by ICP	Turbidity	@RADEDD																																																						
@YMOLD																																																								

Test Description

@2378-TCDD_Dioxin -- 2,3,7,8-TCDD_Dioxin
 @331-PHG -- Perchlorate by EPA 331.0
 @504LOW -- EPA Method 504.1
 @505 -- Organochlorine Pesticides/PCBs
 @515.4 -- Chlorophenoxy Herbicides
 @525 -- Semivolatiles by GCMS
 @531 -- Aldicarb
 @549 -- Diquat and Paraquat
 @551SODA -- EDB/DBCP/HAN by EPA 551.1
 @ANIONS28 -- Chloride, Sulfate by EPA 300.0
 @ANIONS48 -- Nitrate, Nitrite by EPA 300.0
 @DBP_14 -- Chlorite by 300.0
 @DBP_28 -- Disinfection ByProducts by 300.0

Acknowledgement of Samples Received

Addr: **Chameleon Beverage Company**
Chameleon Beverage Company
6444 E 26TH ST
COMMERCE, CA 90040-3214
United States

Attn: Araceli Ramirez
Phone: 3237248223

Client ID: CHAMELEON

Folder #: 997695

Project: ANNUAL

Sample Group: annual-50 state

Project Manager: Rachelle Arada

Phone: 626-386-1106

The following samples were received from you on **April 08, 2022 at 00:36**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
	@HAA -- Haloacetic Acids	
	@ICP -- ICP Metals	
	@ICPMS -- ICPMS Metals	
	@QUANT2000 18HR -- Quantitray Coliforms 18 Hour	
	@RA226 GA -- Radium 226	
	@RA228 GA -- Radium 228	
	@VOA -- Volatile Organics by GCMS	
	@RADEDD -- Gross Alpha/Beta Radiation (Sub)	
	@YMOLD -- Yeast and Mold	

FROM: (323) 724-8223
Fernanda Ortiz
Chameleón Beverage
6444 E. 26th Street
COMMERCE CA 90040

SHIP DATE: 05APR22
ACTWGHT: 35.00 LB
CAD: 5654016IN1ET460

US

BILL SENDER

TO ATT. Rachelle Arada

EEA LABS

750 ROYAL OAKS DR

STE 100

MONROVIA CA 91016

(626) 386-1100


REF: TITLE 21

PO:

DEPT:

(US)

56D.J2/BDF9/FE4A



J221022010501uv


1 of 2

TRK# 7764 9220 9568

MASTER

91016

9622 0019 0 (000 000 0000) 0 00 7764 9220 9568



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: IMPORTANT: TRANSMIT YOUR SHIPPING DATA AND PRINT A MANIFEST:

At the end of each shipping day, you should perform the FedEx Ground End of Day Close procedure to transmit your shipping data to FedEx. To do so, click on the Ground End of Day Close Button. If required, print the pickup manifest that appears. A printed manifest is required to be tendered along with your packages if they are being picked up by FedEx Ground. If you are dropping your packages off at a FedEx drop off location, the manifest is not required.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide and applicable tariff, available upon request. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations, including limitations on our liability, can be found in the current FedEx Service Guide and applicable tariff apply. In no event shall FedEx Ground be liable for any special, incidental, or consequential damages, including, without limitation, loss of profit, loss to the intrinsic value of the package, loss of sale, interest income or attorney's fees. Recovery cannot exceed actual documented loss. Items of extraordinary value are subject to separate limitations of liability set forth in the Service Guide and tariff. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 0971095

SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = NO1A (Observation = 24.5 °C) (Corr. Factor = 0.2 °C) (Final = 24.3 °C)

TYPE OF ICE: Real ☒ Synthetic ☒ No Ice ☒ CONDITION OF ICE: Frozen ☐ Partially Frozen ☐ Thawed ☒ N/A ☒

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: UPS

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤ 6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

VOA and Radon

No Samples with Headspace:

Samples with Headspace (see below):

7) Headspace:

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 515.4, HAA(6251.552), 505, SPME, @CH, 532LCMS, 556, 536, Anatoxin, LCMS methods using 40 ml vials, International clients: None/<6 mm

Samp ID	Bottle #	None/<6 mm	>6mm	Test	Samp ID	Bottle #	None/<6 mm	>6mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors):

RECEIVED BY: <u>Andrew Wilson</u>	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
			Eurofins Eaton Analytical	<u>4-8-22</u>	<u>0330</u>
SAMPLES CHECKED AGAINST COC BY: _____	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
			Eurofins Eaton Analytical		

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Client: Chameleon Beverage Company
Chameleon Beverage Company
6444 E 26TH ST

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Attention: Araceli Ramirez

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Investigation: Analysis per Title 21, Federal Code of Regulations 165.110 - California Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
GROUP I					
PHYSICAL					
Apparent Color (ACU)	SM 2120B	3.0	1	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1.0	1	ND TON	3
pH (Units)	4500HB/E 150	0.10	1	6.7	8.5
Specific Conductance (uS/cm)	SM2510B	2.0	1	11	no std
Total Dissolved Solid (TDS)	SM 2540C	10	1	ND	500
Turbidity (NTU)	EPA 180.1	0.10	1	0.10 NTU	5
GROUP II					
CHEMICAL SUBSTANCE 1					
Milligrams per Liter					
Alkalinity in CaCO3 units	SM 2320B	2.0	1	2.8	no std
Aluminum	EPA 200.8	0.020	1	ND	0.2
Antimony	EPA 200.8	0.0010	1	ND	0.006
Arsenic	EPA 200.8	0.0010	1	ND	0.010
Asbestos by TEM - >10 microns (MFL)	EPA 100.2	0.20	1	ND	no std
Barium	EPA 200.8	0.0020	1	ND	2
Beryllium	EPA 200.8	0.0010	1	ND	0.004
Bicarb Alkalinity as HCO3	SM2330B	2.0	1	3.4	no std
Cadmium	EPA 200.8	0.00050	1	ND	0.005
Calcium	EPA 200.7	1.0	1	ND	no std
Carbonate as CO3	SM2330B	2.0	1	ND	no std
Chloride	EPA 300.0	0.50	1	1.4	250
Chromium	EPA 200.8	0.0010	1	ND	0.1
Copper	EPA 200.8	0.0020	1	ND	1.0
Corrosivity (units)	SM 2330B	-14	1	-4.5	no std
Cyanide	SM 4500CN-F	0.025	1	ND	0.1
Fluoride	SM 4500F-C	0.050	1	ND	1.4
Hydroxide as OH	SM2330B	2.0	1	ND	no std
Iron	EPA 200.7	0.010	1	ND	0.3
Lead	EPA 200.8	0.00050	1	ND	0.005

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_CA

Report: 997695

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
Magnesium	EPA 200.7	0.10	1	ND	no std
Manganese	EPA 200.8	0.0020	1	ND	0.05
Mercury	EPA 200.8	0.00020	1	ND	0.002
Nickel	EPA 200.8	0.0050	1	ND	0.1
Nitrate-N	EPA 300.0	0.050	1	0.086	10
Nitrite-N	EPA 300.0	0.050	1	ND	1
Perchlorate	EPA 331.0	0.00050	1	ND	no std
Phenol	EPA 420.4	0.0010	1	ND	0.001
Potassium	EPA 200.7	1.0	1	ND	no std
Selenium	EPA 200.8	0.0050	1	ND	0.05
Silver	EPA 200.8	0.00050	1	ND	0.1
Sodium	EPA 200.7	1.0	1	2.3	no std
Sulfate	EPA 300.0	0.50	1	ND	250
Surfactants (MBAS)	SM 5540C	0.10	1	ND	no std
Thallium	EPA 200.8	0.0010	1	ND	0.002
Total Hardness as CaCO ₃	EPA 200.7	3.0	1	ND	no std
Total Nitrate + Nitrite	EPA 300.0	0.050	1	0.086	10
Zinc	EPA 200.8	0.020	1	ND	5.0
GROUP III					
CHEMICAL SUBSTANCE 2 (VOC)		Milligrams per Liter			
1,1,1,2-Tetrachloroethane	EPA 524.2	0.00050	1	ND	no std
1,1,1-Trichloroethane	EPA 524.2	0.00050	1	ND	0.20
1,1,2,2-Tetrachloroethane	EPA 524.2	0.00050	1	ND	no std
1,1,2-Trichloroethane	EPA 524.2	0.00050	1	ND	0.005
1,1-Dichloroethane	EPA 524.2	0.00050	1	ND	no std
1,1-Dichloroethene	EPA 524.2	0.00050	1	ND	0.007
1,1-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
1,2,3-Trichlorobenzene	EPA 524.2	0.00050	1	ND	no std
1,2,3-Trichloropropane	EPA 524.2	0.00050	1	ND	no std
1,2,4-Trichlorobenzene	EPA 524.2	0.00050	1	ND	0.07
1,2,4-Trimethylbenzene	EPA 524.2	0.00050	1	ND	no std
1,2-Dichloroethane	EPA 524.2	0.00050	1	ND	0.005
1,2-Dichloropropane	EPA 524.2	0.00050	1	ND	0.005
1,3,5-Trimethylbenzene	EPA 524.2	0.00050	1	ND	no std
1,3-Dichloropropane	EPA 524.2	0.00050	1	ND	no std
2,2-Dichloropropane	EPA 524.2	0.00050	1	ND	no std
Benzene	EPA 524.2	0.00050	1	ND	0.005
Bromobenzene	EPA 524.2	0.00050	1	ND	no std
Bromochloromethane	EPA 524.2	0.00050	1	ND	no std

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_CA
Report: 997695

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
Bromodichloromethane	EPA 524.2	0.00050	1	0.0024	no std
Bromoform	EPA 524.2	0.00050	1	0.0012	no std
Bromomethane	EPA 524.2	0.00050	1	ND	no std
Carbon Tetrachloride	EPA 524.2	0.00050	1	ND	0.005
Chlorobenzene	EPA 524.2	0.00050	1	ND	0.1
Chlorodibromomethane	EPA 524.2	0.00050	1	0.0022	no std
Chloroethane	EPA 524.2	0.00050	1	ND	no std
Chloroform (Trichloromethane)	EPA 524.2	0.00050	1	0.0027	no std
Chloromethane	EPA 524.2	0.00050	1	ND	no std
cis-1,2-Dichloroethylene	EPA 524.2	0.00050	1	ND	0.07
cis-1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Dibromomethane	EPA 524.2	0.00050	1	ND	no std
Dichlorodifluoromethane	EPA 524.2	0.00050	1	ND	no std
Dichloromethane (MeCl ₂)	EPA 524.2	0.00050	1	ND	0.005
Ethyl benzene	EPA 524.2	0.00050	1	ND	0.7
Fluorotrichloromethane-Freon11	EPA 524.2	0.00050	1	ND	no std
Hexachlorobutadiene	EPA 524.2	0.00050	1	ND	no std
Isopropylbenzene	EPA 524.2	0.00050	1	ND	no std
m,p-Xylenes	EPA 524.2	0.00050	1	ND	no std
m-Dichlorobenzene (1,3-DCB)	EPA 524.2	0.00050	1	ND	no std
Methyl ethyl ketone (MEK, Butanone)	EPA 524.2	0.0050	1	ND	no std
MTBE	EPA 524.2	0.00050	1	ND	no std
n-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
n-Propylbenzene	EPA 524.2	0.00050	1	ND	no std
o-Chlorotoluene	EPA 524.2	0.00050	1	ND	no std
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.00050	1	ND	0.6
o-Xylene	EPA 524.2	0.00050	1	ND	no std
p-Chlorotoluene	EPA 524.2	0.00050	1	ND	no std
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.00050	1	ND	0.075
p-Isopropyltoluene	EPA 524.2	0.00050	1	ND	no std
sec-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
Styrene	EPA 524.2	0.00050	1	ND	0.1
tert-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
Tetrachloroethylene (PCE)	EPA 524.2	0.00050	1	ND	0.005
Toluene	EPA 524.2	0.00050	1	ND	1
Total 1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Total THM	EPA 524.2	0.00050	1	0.0085	0.010
Total xylenes	EPA 524.2	0.00050	1	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.00050	1	ND	0.1

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_CA

Report: 997695

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
trans-1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Trichloroethylene (TCE)	EPA 524.2	0.00050	1	ND	0.005
Trichlorotrifluoroethane(F113)	EPA 524.2	0.00050	1	ND	no std
Vinyl chloride (VC)	EPA 524.2	0.00030	1	ND	0.002
GROUP IV					
CHEMICAL SUBSTANCE 3 (NON VOC)		Milligrams per Liter			
1,2,3-Trichloropropane (TCP)	EPA 504.1	0.000020	1	ND	no std
2,3,7,8-TCDD (ug/L)	EPA 1613B	0.0000000050	1	ND	0.00000003
2,4,5-TP (Silvex)	EPA 515.4	0.00020	1	ND	0.05
2,4-D	EPA 515.4	0.00010	1	ND	0.07
3-Hydroxycarbofuran	EPA 531.2	0.00050	1	ND	no std
Alachlor (Alanex)	EPA 505	0.00010	1	ND	0.002
Aldicarb (Temik)	EPA 531.2	0.00050	1	ND	no std
Aldicarb sulfone	EPA 531.2	0.00050	1	ND	no std
Aldicarb sulfoxide	EPA 531.2	0.00050	1	ND	no std
Aldrin	EPA 505	0.000010	1	ND	no std
Atrazine	EPA 525.2	0.000050	1	ND	0.003
Baygon	EPA 531.2	0.00050	1	ND	no std
Bentazon	EPA 515.4	0.00050	1	ND	no std
Benzo(a)pyrene	EPA 525.2	0.000020	1	ND	0.0002
Butachlor	EPA 525.2	0.000050	1	ND	no std
Carbaryl	EPA 531.2	0.00050	1	ND	no std
Carbofuran	EPA 531.2	0.00050	1	ND	0.04
Chlordane	EPA 505	0.00010	1	ND	0.002
Dalapon	EPA 515.4	0.0010	1	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.00060	1	ND	0.4
Di(2-Ethylhexyl)phthalate	EPA 525.2	0.00060	1	ND	0.006
Dibromochloropropane (DBCP)	EPA 551.1	0.000010	1	ND	0.0002
Dibromochloropropane (DBCP)	EPA 504.1	0.000010	1	ND	0.0002
Dicamba	EPA 515.4	0.00010	1	ND	no std
Dieldrin	EPA 505	0.000010	1	ND	no std
Dinoseb	EPA 515.4	0.00020	1	ND	0.007
Diquat	EPA 549.2	0.00040	1	ND	0.02
Endothall	EPA 548.1	0.0050	1	ND	0.1
Endrin	EPA 505	0.000010	1	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.000010	1	ND	0.00005
Ethylene Dibromide (EDB)	EPA 504.1	0.000010	1	ND	0.00005
Glyphosate	EPA 547	0.0060	1	ND	0.7
Heptachlor	EPA 505	0.000010	1	ND	0.0004

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_CA
Report: 997695

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
Heptachlor Epoxide	EPA 505	0.000010	1	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.000050	1	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.000050	1	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.000010	1	ND	0.0002
Methiocarb	EPA 531.2	0.00050	1	ND	no std
Methomyl	EPA 531.2	0.00050	1	ND	no std
Methoxychlor	EPA 505	0.000050	1	ND	0.04
Metolachlor	EPA 525.2	0.000050	1	ND	no std
Metribuzin	EPA 525.2	0.000050	1	ND	no std
Oxamyl (Vydate)	EPA 531.2	0.00050	1	ND	0.2
Paraquat	EPA 549.2	0.0020	1	ND	no std
Pentachlorophenol	EPA 515.4	0.000040	1	ND	0.001
Picloram	EPA 515.4	0.00010	1	ND	0.5
Propachlor	EPA 525.2	0.000050	1	ND	no std
Simazine	EPA 525.2	0.000050	1	ND	0.004
Thiobencarb	EPA 525.2	0.00020	1	ND	no std
Total PCBs	EPA 505	0.00010	1	ND	0.0005
Toxaphene	EPA 505	0.00050	1	ND	0.003
GROUP V					
RADIOACTIVITY					
		Picocuries per Liter			
Gross Alpha	EPA 900.0	2.2	1	<2.18	15
Gross Beta	EPA 900.0	2.0	1	<1.95	50
Radium 226	Ra-226 GA	1.0	1	ND	5
Radium 228	Ra-228 GA	1.0	1	ND	5
Uranium (mg/L)	EPA 200.8	0.0010	1	ND	0.03
GROUP VIa					
BACTERIOLOGICAL					
		Colonies/100 mL			
E Coli Bacteria	SM 9223	1.0	1	<1	1
Total Coliform Bacteria	SM 9223	1.0	1	<1	2.2
GROUP VIb					
BACTERIOLOGICAL-HPC					
		Colony Forming Units per mL			
Heterotrophic Plate Count	SM 9215B	1.0	1	<1	no std
GROUP VII					
Disinfection Byproducts					
		Milligrams per Liter			
Bromate	EPA 317	0.0010	1	ND	0.01
Bromide	EPA 300.0	0.0050	1	ND	no std
Chlorite	EPA 300.0	0.010	1	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.0020	1	ND	0.06

SOQ=Standard of Quality per FDA or California

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_CA
Report: 997695

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
GROUP VIII					
Residual Disinfectants		Milligrams per Liter			
Chloramines	SM 4500CL-G/HACH	0.10	1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.24	1	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.10	1	ND	4

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Client: Chameleon Beverage Company
Chameleon Beverage Company
6444 E 26TH ST

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Attention: Araceli Ramirez

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Investigation: Analysis per Title 21, Federal Code of Regulations 165.110 - New York Limits

ANALYTICAL RESULTS

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
GROUP I					
PHYSICAL					
Apparent Color (ACU)	SM 2120B	3.0	1	ND ACU	15
Odor at 60 C (TON)	SM 2150B	1.0	1	ND TON	3
pH (Units)	4500HB/E 150	0.10	1	6.7	8.5
Specific Conductance (uS/cm)	SM2510B	2.0	1	11	no std
Total Dissolved Solid (TDS)	SM 2540C	10	1	ND	500
Turbidity (NTU)	EPA 180.1	0.10	1	0.10 NTU	5
GROUP II					
CHEMICAL SUBSTANCE 1					
Milligrams per Liter					
Alkalinity in CaCO ₃ units	SM 2320B	2.0	1	2.8	no std
Aluminum	EPA 200.8	0.020	1	ND	0.2
Antimony	EPA 200.8	0.0010	1	ND	0.006
Arsenic	EPA 200.8	0.0010	1	ND	0.010
Asbestos by TEM - >10 microns (MFL)	EPA 100.2	0.20	1	ND	no std
Barium	EPA 200.8	0.0020	1	ND	2
Beryllium	EPA 200.8	0.0010	1	ND	0.004
Bicarb Alkalinity as HCO ₃	SM2330B	2.0	1	3.4	no std
Cadmium	EPA 200.8	0.00050	1	ND	0.005
Calcium	EPA 200.7	1.0	1	ND	no std
Carbonate as CO ₃	SM2330B	2.0	1	ND	no std
Chloride	EPA 300.0	0.50	1	1.4	250
Chromium	EPA 200.8	0.0010	1	ND	0.1
Copper	EPA 200.8	0.0020	1	ND	1.0
Corrosivity (units)	SM 2330B	-14	1	-4.5	no std
Cyanide	SM 4500CN-F	0.025	1	ND	0.1
Fluoride	SM 4500F-C	0.050	1	ND	1.4
Hydroxide as OH	SM2330B	2.0	1	ND	no std
Iron	EPA 200.7	0.010	1	ND	0.3

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_NY
Report: 997695

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
Lead	EPA 200.8	0.00050	1	ND	0.005
Magnesium	EPA 200.7	0.10	1	ND	no std
Manganese	EPA 200.8	0.0020	1	ND	0.05
Mercury	EPA 200.8	0.00020	1	ND	0.002
Nickel	EPA 200.8	0.0050	1	ND	0.1
Nitrate-N	EPA 300.0	0.050	1	0.086	10
Nitrite-N	EPA 300.0	0.050	1	ND	1
Perchlorate	EPA 331.0	0.00050	1	ND	no std
Phenol	EPA 420.4	0.0010	1	ND	0.001
Potassium	EPA 200.7	1.0	1	ND	no std
Selenium	EPA 200.8	0.0050	1	ND	0.05
Silver	EPA 200.8	0.00050	1	ND	0.1
Sodium	EPA 200.7	1.0	1	2.3	no std
Sulfate	EPA 300.0	0.50	1	ND	250
Surfactants (MBAS)	SM 5540C	0.10	1	ND	no std
Thallium	EPA 200.8	0.0010	1	ND	0.002
Total Hardness as CaCO ₃	EPA 200.7	3.0	1	ND	no std
Total Nitrate + Nitrite	EPA 300.0	0.050	1	0.086	10
Zinc	EPA 200.8	0.020	1	ND	5.0
GROUP III					
CHEMICAL SUBSTANCE 2 (VOC)		Milligrams per Liter			
1,1,1,2-Tetrachloroethane	EPA 524.2	0.00050	1	ND	no std
1,1,1-Trichloroethane	EPA 524.2	0.00050	1	ND	0.20
1,1,2,2-Tetrachloroethane	EPA 524.2	0.00050	1	ND	no std
1,1,2-Trichloroethane	EPA 524.2	0.00050	1	ND	0.005
1,1-Dichloroethane	EPA 524.2	0.00050	1	ND	no std
1,1-Dichloroethene	EPA 524.2	0.00050	1	ND	0.007
1,1-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
1,2,3-Trichlorobenzene	EPA 524.2	0.00050	1	ND	no std
1,2,3-Trichloropropane	EPA 524.2	0.00050	1	ND	no std
1,2,4-Trichlorobenzene	EPA 524.2	0.00050	1	ND	0.07
1,2,4-Trimethylbenzene	EPA 524.2	0.00050	1	ND	no std
1,2-Dichloroethane	EPA 524.2	0.00050	1	ND	0.005
1,2-Dichloropropane	EPA 524.2	0.00050	1	ND	0.005
1,3,5-Trimethylbenzene	EPA 524.2	0.00050	1	ND	no std
1,3-Dichloropropane	EPA 524.2	0.00050	1	ND	no std

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_NY
Report: 997695

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
2,2-Dichloropropane	EPA 524.2	0.00050	1	ND	no std
Benzene	EPA 524.2	0.00050	1	ND	0.005
Bromobenzene	EPA 524.2	0.00050	1	ND	no std
Bromochloromethane	EPA 524.2	0.00050	1	ND	no std
Bromodichloromethane	EPA 524.2	0.00050	1	0.0024	no std
Bromoform	EPA 524.2	0.00050	1	0.0012	no std
Bromomethane	EPA 524.2	0.00050	1	ND	no std
Carbon Tetrachloride	EPA 524.2	0.00050	1	ND	0.005
Chlorobenzene	EPA 524.2	0.00050	1	ND	0.1
Chlorodibromomethane	EPA 524.2	0.00050	1	0.0022	no std
Chloroethane	EPA 524.2	0.00050	1	ND	no std
Chloroform (Trichloromethane)	EPA 524.2	0.00050	1	0.0027	no std
Chloromethane	EPA 524.2	0.00050	1	ND	no std
cis-1,2-Dichloroethylene	EPA 524.2	0.00050	1	ND	0.07
cis-1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Dibromomethane	EPA 524.2	0.00050	1	ND	no std
Dichlorodifluoromethane	EPA 524.2	0.00050	1	ND	no std
Dichloromethane (MeCl ₂)	EPA 524.2	0.00050	1	ND	0.005
Ethyl benzene	EPA 524.2	0.00050	1	ND	0.7
Fluorotrichloromethane-Freon11	EPA 524.2	0.00050	1	ND	no std
Hexachlorobutadiene	EPA 524.2	0.00050	1	ND	no std
Isopropylbenzene	EPA 524.2	0.00050	1	ND	no std
m,p-Xylenes	EPA 524.2	0.00050	1	ND	no std
m-Dichlorobenzene (1,3-DCB)	EPA 524.2	0.00050	1	ND	no std
Methyl ethyl ketone (MEK, Butanone)	EPA 524.2	0.0050	1	ND	no std
MTBE	EPA 524.2	0.00050	1	ND	no std
n-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
n-Propylbenzene	EPA 524.2	0.00050	1	ND	no std
o-Chlorotoluene	EPA 524.2	0.00050	1	ND	no std
o-Dichlorobenzene (1,2-DCB)	EPA 524.2	0.00050	1	ND	0.6
o-Xylene	EPA 524.2	0.00050	1	ND	no std
p-Chlorotoluene	EPA 524.2	0.00050	1	ND	no std
p-Dichlorobenzene (1,4-DCB)	EPA 524.2	0.00050	1	ND	0.075
p-Isopropyltoluene	EPA 524.2	0.00050	1	ND	no std
sec-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
Styrene	EPA 524.2	0.00050	1	ND	0.1

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_NY
Report: 997695

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
tert-Butylbenzene	EPA 524.2	0.00050	1	ND	no std
Tetrachloroethylene (PCE)	EPA 524.2	0.00050	1	ND	0.005
Toluene	EPA 524.2	0.00050	1	ND	1
Total 1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Total THM	EPA 524.2	0.00050	1	0.0085	0.080
Total xylenes	EPA 524.2	0.00050	1	ND	10
trans-1,2-Dichloroethylene	EPA 524.2	0.00050	1	ND	0.1
trans-1,3-Dichloropropene	EPA 524.2	0.00050	1	ND	no std
Trichloroethylene (TCE)	EPA 524.2	0.00050	1	ND	0.005
Trichlorotrifluoroethane(F113)	EPA 524.2	0.00050	1	ND	no std
Vinyl chloride (VC)	EPA 524.2	0.00030	1	ND	0.002
GROUP IV					
CHEMICAL SUBSTANCE 3 (NON VOC)		Milligrams per Liter			
1,2,3-Trichloropropane (TCP)	EPA 504.1	0.000020	1	ND	no std
2,3,7,8-TCDD (ug/L)	EPA 1613B	0.0000000050	1	ND	0.00000003
2,4,5-TP (Silvex)	EPA 515.4	0.00020	1	ND	0.05
2,4-D	EPA 515.4	0.00010	1	ND	0.07
3-Hydroxycarbofuran	EPA 531.2	0.00050	1	ND	no std
Alachlor (Alanex)	EPA 505	0.00010	1	ND	0.002
Aldicarb (Temik)	EPA 531.2	0.00050	1	ND	no std
Aldicarb sulfone	EPA 531.2	0.00050	1	ND	no std
Aldicarb sulfoxide	EPA 531.2	0.00050	1	ND	no std
Aldrin	EPA 505	0.000010	1	ND	no std
Atrazine	EPA 525.2	0.000050	1	ND	0.003
Baygon	EPA 531.2	0.00050	1	ND	no std
Bentazon	EPA 515.4	0.00050	1	ND	no std
Benzo(a)pyrene	EPA 525.2	0.000020	1	ND	0.0002
Butachlor	EPA 525.2	0.000050	1	ND	no std
Carbaryl	EPA 531.2	0.00050	1	ND	no std
Carbofuran	EPA 531.2	0.00050	1	ND	0.04
Chlordane	EPA 505	0.00010	1	ND	0.002
Dalapon	EPA 515.4	0.0010	1	ND	0.2
Di-(2-Ethylhexyl)adipate	EPA 525.2	0.00060	1	ND	0.4
Di(2-Ethylhexyl)phthalate	EPA 525.2	0.00060	1	ND	0.006
Dibromochloropropane (DBCP)	EPA 551.1	0.000010	1	ND	0.0002
Dibromochloropropane (DBCP)	EPA 504.1	0.000010	1	ND	0.0002

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_NY
Report: 997695

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
Dicamba	EPA 515.4	0.00010	1	ND	no std
Dieldrin	EPA 505	0.000010	1	ND	no std
Dinoseb	EPA 515.4	0.00020	1	ND	0.007
Diquat	EPA 549.2	0.00040	1	ND	0.02
Endothall	EPA 548.1	0.0050	1	ND	0.1
Endrin	EPA 505	0.000010	1	ND	0.002
Ethylene Dibromide (EDB)	EPA 551.1	0.000010	1	ND	0.00005
Ethylene Dibromide (EDB)	EPA 504.1	0.000010	1	ND	0.00005
Glyphosate	EPA 547	0.0060	1	ND	0.7
Heptachlor	EPA 505	0.000010	1	ND	0.0004
Heptachlor Epoxide	EPA 505	0.000010	1	ND	0.0002
Hexachlorobenzene	EPA 525.2	0.000050	1	ND	0.001
Hexachlorocyclopentadiene	EPA 525.2	0.000050	1	ND	0.05
Lindane (gamma-BHC)	EPA 505	0.000010	1	ND	0.0002
Methiocarb	EPA 531.2	0.00050	1	ND	no std
Methomyl	EPA 531.2	0.00050	1	ND	no std
Methoxychlor	EPA 505	0.000050	1	ND	0.04
Metolachlor	EPA 525.2	0.000050	1	ND	no std
Metribuzin	EPA 525.2	0.000050	1	ND	no std
Oxamyl (Vydate)	EPA 531.2	0.00050	1	ND	0.2
Paraquat	EPA 549.2	0.0020	1	ND	no std
Pentachlorophenol	EPA 515.4	0.000040	1	ND	0.001
Picloram	EPA 515.4	0.00010	1	ND	0.5
Propachlor	EPA 525.2	0.000050	1	ND	no std
Simazine	EPA 525.2	0.000050	1	ND	0.004
Thiobencarb	EPA 525.2	0.00020	1	ND	no std
Total PCBs	EPA 505	0.00010	1	ND	0.0005
Toxaphene	EPA 505	0.00050	1	ND	0.003

GROUP V

RADIOACTIVITY

Picocuries per Liter

Gross Alpha	EPA 900.0	2.2	1	<2.18	15
Gross Beta	EPA 900.0	2.0	1	<1.95	50
Radium 226	Ra-226 GA	1.0	1	ND	5
Radium 228	Ra-228 GA	1.0	1	ND	5
Uranium (mg/L)	EPA 200.8	0.0010	1	ND	0.03

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory FDA_NY
Report: 997695

Sample Id: Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL
Date Sampled: 4/8/2022

Report Date: 07/20/2022
Date Received: 04/08/2022
Sample No: 202204080117

Parameter	Method	Reporting Limit	Dilution	Result	SOQ
GROUP VIa					
BACTERIOLOGICAL		Colonies/100 mL			
E Coli Bacteria	SM 9223	1.0	1	<1	1
Total Coliform Bacteria	SM 9223	1.0	1	<1	2.2
GROUP VIb					
BACTERIOLOGICAL-HPC		Colony Forming Units per mL			
Heterotrophic Plate Count	SM 9215B	1.0	1	<1	no std
GROUP VII					
Disinfection Byproducts		Milligrams per Liter			
Bromate	EPA 317	0.0010	1	ND	0.01
Bromide	EPA 300.0	0.0050	1	ND	no std
Chlorite	EPA 300.0	0.010	1	ND	1
D/DBP Haloacetic Acids (HAA5)	SM 6251B	0.0020	1	ND	0.06
GROUP VIII					
Residual Disinfectants		Milligrams per Liter			
Chloramines	SM 4500CL-G/HACH	0.10	1	ND	4
Chlorine Dioxide	SM 4500CLO2-D/HACH	0.24	1	ND	0.8
Total Chlorine Residual	SM 4500CL-G/HACH	0.10	1	ND	4

SOQ=Standard of Quality per FDA or New York

ND=Not detected at the specified limit

All results reported in milligrams per liter unless otherwise noted

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 997695
Project: ANNUAL
Group: annual-50 state

Chameleon Beverage Company
Araceli Ramirez
Chameleon Beverage Company
6444 E 26TH ST
COMMERCE, CA 90040-3214
United States

Client Specific Comments

Report format not TNI (The NELAC Institute) Standard compliant, but is in client specific format.

Folder Comments

Analytical results for Gross Alpha and Beta are submitted by Pace Analytical Services, LLC

July 06, 2022

Ms. Jaclyn L. Contreras
Eurofins Eaton Analytical, LLC
750 Royal Oaks Drive, 100
Monrovia, CA 910163629

RE: Project: PACE PA 997695
Pace Project No.: 30493162

Dear Ms. Contreras:

Enclosed are the analytical results for sample(s) received by the laboratory on May 26, 2022. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Greensburg

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Skyler C. Richmond
skyler.richmond@pacelabs.com
(724)850-5600
Project Manager

Enclosures

cc: Yolanda Martin, Eurofins Eaton Analytical, LLC
Mr. Aleksandar D. Tomovich, Eurofins Eaton Analytical, LLC



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: PACE PA 997695

Pace Project No.: 30493162

Pace Analytical Services Pennsylvania

1638 Roseytown Rd Suites 2,3&4, Greensburg, PA 15601

ANAB DOD-ELAP Rad Accreditation #: L2417

Alabama Certification #: 41590

Arizona Certification #: AZ0734

Arkansas Certification

California Certification #: 04222CA

Colorado Certification #: PA01547

Connecticut Certification #: PH-0694

Delaware Certification

EPA Region 4 DW Rad

Florida/TNI Certification #: E87683

Georgia Certification #: C040

Florida: Cert E871149 SEKS WET

Guam Certification

Hawaii Certification

Idaho Certification

Illinois Certification

Indiana Certification

Iowa Certification #: 391

Kansas/TNI Certification #: E-10358

Kentucky Certification #: KY90133

KY WW Permit #: KY0098221

KY WW Permit #: KY0000221

Louisiana DHH/TNI Certification #: LA180012

Louisiana DEQ/TNI Certification #: 4086

Maine Certification #: 2017020

Maryland Certification #: 308

Massachusetts Certification #: M-PA1457

Michigan/PADEP Certification #: 9991

Missouri Certification #: 235

Montana Certification #: Cert0082

Nebraska Certification #: NE-OS-29-14

Nevada Certification #: PA014572018-1

New Hampshire/TNI Certification #: 297617

New Jersey/TNI Certification #: PA051

New Mexico Certification #: PA01457

New York/TNI Certification #: 10888

North Carolina Certification #: 42706

North Dakota Certification #: R-190

Ohio EPA Rad Approval: #41249

Oregon/TNI Certification #: PA200002-010

Pennsylvania/TNI Certification #: 65-00282

Puerto Rico Certification #: PA01457

Rhode Island Certification #: 65-00282

South Dakota Certification

Tennessee Certification #: 02867

Texas/TNI Certification #: T104704188-17-3

Utah/TNI Certification #: PA014572017-9

USDA Soil Permit #: P330-17-00091

Vermont Dept. of Health: ID# VT-0282

Virgin Island/PADEP Certification

Virginia/VELAP Certification #: 460198

Washington Certification #: C868

West Virginia DEP Certification #: 143

West Virginia DHHR Certification #: 9964C

Wisconsin Approve List for Rad

Wyoming Certification #: 8TMS-L

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: PACE PA 997695

Pace Project No.: 30493162

Lab ID	Sample ID	Matrix	Date Collected	Date Received
30493162001	202204080117	Drinking Water	04/08/22 10:10	05/26/22 09:30

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: PACE PA 997695

Pace Project No.: 30493162

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
30493162001	202204080117	EPA 900.0	ERT	2	PASI-PA

PASI-PA = Pace Analytical Services - Greensburg

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

PROJECT NARRATIVE

Project: PACE PA 997695

Pace Project No.: 30493162

Method: EPA 900.0

Description: 900.0 Gross Alpha/Beta

Client: Eurofins Eaton Analytical

Date: July 06, 2022

General Information:

1 sample was analyzed for EPA 900.0 by Pace Analytical Services Greensburg. All samples were received in acceptable condition with any exceptions noted below or on the chain-of custody and/or the sample condition upon receipt form (SCUR) attached at the end of this report.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank, where applicable, with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Additional Comments:

This data package has been reviewed for quality and completeness and is approved for release.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS - RADIOCHEMISTRY

Project: PACE PA 997695

Pace Project No.: 30493162

Sample: 202204080117 **Lab ID:** 30493162001 Collected: 04/08/22 10:10 Received: 05/26/22 09:30 Matrix: Drinking Water

PWS: **Site ID:** **Sample Type:**

Comments: • Preservative not listed on COC.
 • Sampler's name and signature not on COC.

Parameters	Method	Act ± Unc (MDC) Carr Trac	Units	Analyzed	CAS No.	Qual
Pace Analytical Services - Greensburg						
Gross Alpha	EPA 900.0	-0.333 ± 0.860 (2.18) C:NA T:NA	pCi/L	06/29/22 19:52	12587-46-1	
Gross Beta	EPA 900.0	-0.455 ± 0.808 (1.95) C:NA T:NA	pCi/L	06/29/22 19:52	12587-47-2	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL - RADIOCHEMISTRY

Project: PACE PA 997695

Pace Project No.: 30493162

QC Batch: 513678

Analysis Method: EPA 900.0

QC Batch Method: EPA 900.0

Analysis Description: 900.0 Gross Alpha/Beta

Laboratory: Pace Analytical Services - Greensburg

Associated Lab Samples: 30493162001

METHOD BLANK: 2489590

Matrix: Water

Associated Lab Samples: 30493162001

Parameter	Act ± Unc (MDC) Carr Trac	Units	Analyzed	Qualifiers
Gross Alpha	-0.545 ± 0.512 (1.85) C:NA T:NA	pCi/L	06/30/22 08:52	
Gross Beta	0.298 ± 0.685 (1.63) C:NA T:NA	pCi/L	06/30/22 08:52	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: PACE PA 997695

Pace Project No.: 30493162

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Act - Activity

Unc - Uncertainty: For Safe Drinking Water Act (SDWA) analyses, the reported Unc. is the calculated Count Uncertainty (95% confidence interval) using a coverage factor of 1.96. For all other matrices (non-SDWA), the reported Unc. is the calculated Expanded Uncertainty (aka Combined Standard Uncertainty, CSU), reported at the 95% confidence interval using a coverage factor of 1.96.

Gamma Spec: The Unc. reported for all gamma-spectroscopy analyses (EPA 901.1), is the calculated Expanded Uncertainty (CSU) at the 95.4% confidence interval, using a coverage factor of 2.0.

(MDC) - Minimum Detectable Concentration

Trac - Tracer Recovery (%)

Carr - Carrier Recovery (%)

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Ship To:

PACE Analytical Services, Inc.
1638 Roseytown Road, Suite 2
Suite 2
Greensburg, PA 15601

Phone: 724-850-5600 Fax: 724-850-5601

Folder #: 997695
Report Due: 05/06/2022

Reports: Jackie Contreras Sub-Contracting Administrator
EMAIL TO: Eaton-MonroviaSubContract@eurofins.com
Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016
Phone (626) 386-1165 Fax (626) 386-1122
Invoices to: Eurofins Eaton Analytical, LLC
Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the
Specified State Certification # and
Exp Date for requested tests + matrix.

Samples from: CALIFORNIA

Sample ID 202204080117 Client Sample ID for reference onl
Product CBC 22 095 0741 EXP APR 05 24 P PURIFIED WATER 500mL

Sample type: Sample Event: Sample Date & Time 04/08/22 1010 DW Clip Code PWSID

Method EPA 900.0 Prep Method Analysis Requested Gross Alpha/Beta Radiation (Sub) Sample Point ID: Static ID: JLS

WO#: 30493162



30493162

Relinquished by: [Signature] Sample Control Date 05/24/22 Time 9:26
Received by: [Signature] Date 5-26-22 Time 9:30
Relinquished by: Sample Control Date Time
Received by: Sample Control Date Time

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS
An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Pittsburgh Lab Sample Condition Upon Receipt

Client Name: Eurofins

Project # _____

Courier: ☒ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☐ Pace Other _____Tracking #: 5717 5704 9429

Label	<u>PS</u>
LIMS Login	<u>VP</u>

Custody Seal on Cooler/Box Present: ☐ yes ☒ no Seals intact: ☐ yes ☒ noThermometer Used _____ Type of Ice: Wet Blue None

Cooler Temperature _____ Observed Temp _____ °C Correction Factor: _____ °C Final Temp: _____ °C

Temp should be above freezing to 6°C

pH paper Lot#

1004611Date and Initials of person examining contents: PS 5/31/22 11:05

Comments:

Yes No N/A

Chain of Custody Present: ☒ Yes ☐ No ☐ N/AChain of Custody Filled Out: ☒ Yes ☐ No ☐ N/AChain of Custody Relinquished: ☒ Yes ☐ No ☐ N/ASampler Name & Signature on COC: ☐ Yes ☒ No ☐ N/ASample Labels match COC: ☒ Yes ☐ No ☐ N/A

-Includes date/time/ID

Matrix: DWSamples Arrived within Hold Time: ☒ Yes ☐ No ☐ N/AShort Hold Time Analysis (<72hr remaining): ☐ Yes ☒ No ☐ N/ARush Turn Around Time Requested: ☐ Yes ☒ No ☐ N/ASufficient Volume: ☒ Yes ☐ No ☐ N/ACorrect Containers Used: ☒ Yes ☐ No ☐ N/A-Pace Containers Used: ☐ Yes ☒ No ☐ N/AContainers Intact: ☒ Yes ☐ No ☐ N/AOrthophosphate field filtered: ☐ Yes ☐ No ☒ N/AHex Cr Aqueous sample field filtered: ☐ Yes ☐ No ☒ N/AOrganic Samples checked for dechlorination: ☐ Yes ☐ No ☒ N/AFiltered volume received for Dissolved tests: ☐ Yes ☐ No ☒ N/AAll containers have been checked for preservation: ☒ Yes ☐ No ☐ N/A

exceptions: VOA, coliform, TOC, O&G, Phenolics, Radon, Non-aqueous matrix

All containers meet method preservation requirements: ☐ Yes ☒ No ☐ N/A

Initial when completed

PS

Date/time of preservation

5/31/22

Lot # of added preservative

HNO₃ DL22-0473Headspace in VOA Vials (>6mm): ☐ Yes ☐ No ☒ N/ATrip Blank Present: ☐ Yes ☐ No ☒ N/ATrip Blank Custody Seals Present: ☐ Yes ☐ No ☒ N/ARad Samples Screened < 0.5 mrem/hr: ☒ Yes ☐ No ☐ N/A

Initial when completed

PSDate: 5/31/22Survey Meter SN: 1563

Client Notification/ Resolution:

Person Contacted: _____ Date/Time: _____ Contacted By: _____

Comments/ Resolution: _____

☐ A check in this box indicates that additional information has been stored in ereports.

Note: Whenever there is a discrepancy affecting North Carolina compliance samples, a copy of this form will be sent to the North Carolina DEHNR Certification Office (i.e. out of hold, incorrect preservative, out of temp, incorrect containers)

*PM review is documented electronically in LIMS. When the Project Manager closes the SRF Review schedule in LIMS. The review is in the Status section of the Workorder Edit Screen.

W0#: 30493162

PM: SCR Due Date: 06/17/22
CLIENT: EUROF_EATON

Profile Number

PALE PA 997695

Notes

WO#: 30493162

PM: SCR

CLIENT: EUROF EATON

Container Codes


Plastic / Misc.

GCUB	1 Gallon Cubitainer
12GN	1/2 Gallon Cubitainer
SP5T	120mL Coliform Na Thiosulfate
BP1N	1L plastic HNO3
BP1U	1L plastic unpreserved
BP3S	250mL plastic H2SO4
BP3N	250mL plastic HNO3
BP3U	250mL plastic unpreserved
BP3C	250ml plastic NAOH
BP2S	500mL plastic H2SO4
BP2U	500mL plastic unpreserved

EZI	5g Encore
VOAK	Kit for Volatile Solid
I	Wipe/Swab
ZPLC	Ziploc Bag

WT	Water
SL	Solid
OL	Non-aqueous liquid
WP	Wipe

Quality Control Sample Performance Assessment



Test: Gross Alpha and Gross Beta
Analyst: ERT
Date: 6/27/2022
Batch ID: 67409
Matrix: DW

Method Blank Assessment	Gross Alpha	Gross Beta
MB Sample ID	2489590	
MB concentration:	-0.545	0.298
MB 2 Sigma CSU:	0.512	0.685
MB MDC:	1.848	1.633
MB Numerical Performance Indicator:	-2.09	0.85
MB Status vs. MDC:	NA	NA
	Pass	Pass

Laboratory Control Sample Assessment	LCS#	Y or N?	N	Gross Alpha	Gross Beta
Count Date:	7/1/2022				
Spike ID:	21-047				
Spike Concentration (pCi/mL):	29.847				
Volume Used (mL):	0.100				
Aliquot Volume (L, g, F):	0.201				
Target Conc. (pCi/L, g, F):	14.818				
Uncertainty (Calculated):	0.267				
Result (pCi/L, g, F):	15.021				
LCS/LCSD 2 Sigma CU (pCi/L, g, F):	3.047				
Numerical Performance Indicator:	0.13				
Percent Recovery:	101.37%				
Status vs Numerical Indicator:	NA				
Upper % Recovery Limits:	121.00%				
Lower % Recovery Limits:	69.00%				

Duplicate Sample Assessment	Sample ID:	Enter Duplicate sample IDs if other than LCS/LCSD in the space below.
Duplicate Sample ID:	30494073001	
Sample Result (pCi/L, g, F):	0.383	4.683
Sample Duplicate Result (pCi/L, g, F):	0.730	1.228
Sample Duplicate Result (pCi/L, g, F):	0.102	5.366
Sample Duplicate Result 2 Sigma CU (pCi/L, g, F):	0.783	1.277
Are sample and/or duplicate results below MDC?	See Below ##	
Duplicate Numerical Performance Indicator:	0.515	-0.745
Duplicate RPD:	118.07%	13.39%
Duplicate Status vs Numerical Indicator:	Pass	Pass
Duplicate Status vs RPD:	Fail***	Pass
% RPD Limit:	34.00%	29.00%

Evaluation of duplicate precision is not applicable if either the sample or duplicate results are below the MDC.

Comments:

***Batch must be re-prepped due to unacceptable precision.

Handwritten signature and date 7/5/22

Analyst Must Manually Enter All Fields Highlighted in Yellow.

Sample Matrix Spike Control Assessment	Gross Alpha	Gross Beta	Gross Alpha	Gross Beta
Sample Collection Date:	6/2/2022			
Sample I.D.	30494073001			
Sample MS I.D.	30494073001MS			
MS/MSD Decay Corrected Spike Concentration (pCi/mL):	21-047	21-015GB		
Spike Volume Used in MS (mL):	29.847	64.537		
Spike Volume Used in MSD (mL):	0.20	0.20		
MS Aliquot (L, g, F):	0.202	0.202		
MS Target Conc. (pCi/L, g, F):	29.601	64.006		
MSD Aliquot (L, g, F):				
MSD Target Conc. (pCi/L, g, F):				
Spike uncertainty (calculated):	0.533	0.307		
Spike Duplicate uncertainty (calculated):				
Sample Result:	0.383	4.683		
Sample Result 2 Sigma CU (pCi/L, g, F):	0.730	1.228		
Sample Matrix Spike Result:	32.152	68.716		
Sample Matrix Spike Duplicate Result:	4.325	4.038		
MS Numerical Performance Indicator:	0.953	0.008		
MS Percent Recovery:	107.26%	100.03%		
MSD Percent Recovery:				
MS Status vs Numerical Indicator:	Pass	Pass		
MSD Status vs Numerical Indicator:	Pass	Pass		
MS/MSD Upper % Recovery Limits:	135.00%	130.00%		
MS/MSD Lower % Recovery Limits:	55.00%	79.00%		

Matrix Spike/Matrix Spike Duplicate Sample Assessment	Sample ID:	Sample MS I.D.
Sample I.D.	30494073001	
Sample MS I.D.	30494073001MS	
Sample Matrix Spike Result:		
Matrix Spike Result 2 Sigma CU (pCi/L, g, F):		
Sample Matrix Spike Duplicate Result:		
Matrix Spike Duplicate Result 2 Sigma CU (pCi/L, g, F):		
Duplicate Numerical Performance Indicator:		
(Based on the MS / MSD Recoveries) MS / MSD Duplicate RPD:		
MS / MSD Duplicate Status vs Numerical Indicator:		
MS / MSD Duplicate Status vs RPD:		
% RPD Limit:		